



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Sohil

Run Number: LB120800

Slope : 99.8

BalanceID: WC SC-4

pH Meter ID : WC pH meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2905
BUFFER PH 7.00 GREEN 1PINT PK6	W2906
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.9	4.01	06/23/2022	09:00
2	CAL2	1	Water	NA	NA	21.8	7.00	06/23/2022	09:02
3	CAL3	1	Water	NA	NA	20.7	10.05	06/23/2022	09:04
4	ICV	1	Water	NA	NA	21.8	6.99	06/23/2022	09:05
5	CCV1	1	Water	NA	NA	20.9	2.02	06/23/2022	09:07
6	N3450-01	1	Solid	20.02	20	20.2	7.66	06/23/2022	09:10
7	N3450-01DUP	1	Solid	20.03	20	20.3	7.67	06/23/2022	09:12
8	N3450-02	1	Solid	20.04	20	20.2	7.54	06/23/2022	09:14
9	N3450-03	1	Solid	20.04	20	20.4	7.40	06/23/2022	09:16
10	N3450-04	1	Solid	20.03	20	20.2	7.46	06/23/2022	09:18
11	CCV2	1	Water	NA	NA	20.3	12.03	06/23/2022	09:20

WORKLIST(Hardcopy Internal Chain)

LB120800

Worklist Name : ph-062322

Worklist ID : 160797

Department : Wet-Chemistry

Date : 06-23-2022 08:29:09

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
06/22/2022	Solid	N3450-01	pH	Cool 4 deg C	CORN02	D11	PH1-BOT-16	06/22/2022	9045D
06/22/2022	Solid	N3450-02	pH	Cool 4 deg C	CORN02	D11	PH1-BOT-17	06/22/2022	9045D
06/22/2022	Solid	N3450-03	pH	Cool 4 deg C	CORN02	D11	PH1-BOT-18	06/22/2022	9045D
06/22/2022	Solid	N3450-04	pH	Cool 4 deg C	CORN02	D11	PH1-BOT-19	06/22/2022	9045D
06/22/2022	Solid	N3468-01	pH	Cool 4 deg C	EART12	D11	XRDS-SCREENFILL-CLEAN-FILL-SP-A-COMP	06/22/2022	9041A
06/22/2022	Solid	N3469-01	pH	Cool 4 deg C	EART12	D11	XRDS-SCREENFILL-CLEAN-FILL-SP-1-COMP	06/22/2022	9041A

Date/Time 6/23/22 8:45
 Raw Sample Received by: JS (uc)
 Raw Sample Relinquished by: JS (uc)

Date/Time 6/23/22 9:30
 Raw Sample Received by: JS (uc)
 Raw Sample Relinquished by: JS (uc)