

LB120801


7471B

INSTRUMENT ID: CVL

Linear

 μ Abs.:

53372



A= 0.0000e+000

B= 1.8738e-004

C= 1.9298e-002

Rho= 0.9999882

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	7vp
0.0	0.000	0.017	0.017	-12	0.000	-12					
0.2	0.200	0.211	0.011	1023	0.0 %	1023					
2.5	2.500	2.470	-0.030	13081	0.0 %	13081					
5.0	5.000	4.997	-0.003	26564	0.0 %	26564					
7.5	7.500	7.485	-0.015	39841	0.0 %	39841					
10.0	10.000	10.020	0.020	53372	0.0 %	53372					

7vp
1
5
-1
0
0
0

LB120801 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 23 Jun 2022 14:15:34

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-12	-	0.0000	7471B	PPB	23 Jun 2022 14:19:44	S	Std
0.2 - 1	50.2	1023	-	0.2000	7471B	PPB	23 Jun 2022 14:22:01	S	Std
2.5 - 1	52.5	13081	-	2.5000	7471B	PPB	23 Jun 2022 14:24:18	S	Std
5.0 - 1	55	26564	-	5.0000	7471B	PPB	23 Jun 2022 14:26:35	S	Std
7.5 - 1	57.5	39841	-	7.5000	7471B	PPB	23 Jun 2022 14:28:52	S	Std
10.0 - 1	510	53372	-	10.0000	7471B	PPB	23 Jun 2022 14:31:08	S	Std
ICV07 - 1	ICV07	19594	3.6908	-	7471B	PPB	23 Jun 2022 14:33:52	U	SMPL
ICB07 - 1	ICB07	-77	0.0049	-	7471B	PPB	23 Jun 2022 14:36:06	U	SMPL
CCV25 - 1	CCV25	26633	5.0098	-	7471B	PPB	23 Jun 2022 14:38:22	U	SMPL
CCB25 - 1	CCB25	-97	0.0011	-	7471B	PPB	23 Jun 2022 14:40:38	U	SMPL
CRA - 1	CRA	1020	0.2104	-	7471B	PPB	23 Jun 2022 14:42:54	U	SMPL
HighStd - 1	HighStd	53192	9.9864	-	7471B	PPB	23 Jun 2022 14:45:10	U	SMPL
ChkStd - 1	ChkStd	37346	7.0171	-	7471B	PPB	23 Jun 2022 14:47:25	U	SMPL
PB145769BL - 1	PBS	-214	-0.0208	-	7471B	PPB	23 Jun 2022 14:49:43	U	SMPL
PB145769BS - 1	LCSS	19713	3.7131	-	7471B	PPB	23 Jun 2022 14:55:51	U	SMPL
N3430-01 - 1	SB-4-0-2-20220621	713	0.1529	-	7471B	PPB	23 Jun 2022 14:58:08	U	SMPL
N3430-03 - 1	SB-5-0-2-20220621	36632	6.8834	-	7471B	PPB	23 Jun 2022 15:00:24	U	SMPL
N3430-04 - 1	SB-5-6-8-20220621	-26	0.0144	-	7471B	PPB	23 Jun 2022 15:02:42	U	SMPL
N3431-01 - 1	WC-14B-2-2	63411	11.9012	-	7471B	PPB	23 Jun 2022 15:05:01	U	SMPL
N3431-07 - 1	WC-14B-2-3	30795	5.7896	-	7471B	PPB	23 Jun 2022 15:07:16	U	SMPL
CCV26 - 1	CCV26	26080	4.9061	-	7471B	PPB	23 Jun 2022 15:09:34	U	SMPL
CCB26 - 1	CCB26	-104	-0.0002	-	7471B	PPB	23 Jun 2022 15:11:51	U	SMPL
N3431-13 - 1	WC-14B-2-4	30694	5.7707	-	7471B	PPB	23 Jun 2022 15:14:07	U	SMPL
N3431-19 - 1	WC-14B-2-5	34590	6.5007	-	7471B	PPB	23 Jun 2022 15:16:22	U	SMPL
N3450-01 - 1	PH1-BOT-16	977	0.2024	-	7471B	PPB	23 Jun 2022 15:18:40	U	SMPL
N3450-02 - 1	PH1-BOT-17	989	0.2046	-	7471B	PPB	23 Jun 2022 15:20:57	U	SMPL
N3450-03 - 1	PH1-BOT-18	842	0.1771	-	7471B	PPB	23 Jun 2022 15:23:12	U	SMPL
N3450-04 - 1	PH1-BOT-19	736	0.1572	-	7471B	PPB	23 Jun 2022 15:25:28	U	SMPL
N3450-04DUP - 1	PH1-BOT-19DUP	727	0.1555	-	7471B	PPB	23 Jun 2022 15:27:45	U	SMPL
N3450-04MS - 1	PH1-BOT-19MS	21092	3.9715	-	7471B	PPB	23 Jun 2022 15:32:10	U	SMPL
N3450-04MSD - 1	PH1-BOT-19MSD	20498	3.8602	-	7471B	PPB	23 Jun 2022 15:34:27	U	SMPL
CCV27 - 1	CCV27	26618	5.0069	-	7471B	PPB	23 Jun 2022 15:36:43	U	SMPL
CCB27 - 1	CCB27	-102	0.0002	-	7471B	PPB	23 Jun 2022 15:39:00	U	SMPL
N3450-04LX5 - 1		33	0.0255	-	7471B	PPB	23 Jun 2022 15:41:17	U	SMPL
N3450-04A - 1		20587	3.8769	-	7471B	PPB	23 Jun 2022 15:43:33	U	SMPL
N3431-01DLX2 - 1		31615	5.9433	-	7471B	PPB	23 Jun 2022 15:45:48	U	SMPL
CCV28 - 1	CCV28	26632	5.0096	-	7471B	PPB	23 Jun 2022 15:50:58	U	SMPL
CCB28 - 1	CCB28	-82	0.0039	-	7471B	PPB	23 Jun 2022 15:53:13	U	SMPL