



Analytical Summary Report

Analysis Method: 9045D

Parameter: Corrosivity

Run Number: LB120805

BalanceID:

Analyst By : Guerda

Supervisor Review By : mohammad

Slope : 99.8

pH Meter ID : WC pH meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2905
BUFFER PH 7.00 GREEN 1PINT PK6	W2906
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	06/23/2022	10:00
2	CAL2	1	Water	NA	NA	20.7	7.00	06/23/2022	10:04
3	CAL3	1	Water	NA	NA	20.7	10.05	06/23/2022	10:06
4	ICV	1	Water	NA	NA	21.8	6.99	06/23/2022	10:08
5	CCV1	1	Water	NA	NA	20.9	2.01	06/23/2022	10:10
6	N3384-02	1	Solid	20	20	19.2	5.80	06/23/2022	10:12
7	N3384-02DUP	1	Solid	20	20	19.2	5.80	06/23/2022	10:14
8	N3422-02	1	Solid	20	20	19.9	5.31	06/23/2022	10:16
9	N3422-08	1	Solid	20	20	19.2	5.11	06/23/2022	10:16
10	N3431-02	1	Solid	20	20	19.4	6.04	06/23/2022	10:20
11	N3431-08	1	Solid	20	20	19.4	6.74	06/23/2022	10:22
12	N3431-14	1	Solid	20	20	19.1	7.23	06/23/2022	10:22
13	N3431-20	1	Solid	20	20	19.0	7.14	06/23/2022	10:24
14	N3447-01	1	Solid	20	20	20.7	5.00	06/23/2022	10:26
15	CCV2	1	Water	NA	NA	20.3	12.03	06/23/2022	10:30

WORKLIST(Hardcopy Internal Chain)

LB120805

Worklist Name : CORROSIVITY-061722 Worklist ID : 160644 Department : Wet-Chemistry Date : 06-17-2022 12:01:00

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
06/15/2022	Solid	N3384-02	B Corrosivity	Cool 4 deg C	DVIR01	D11	WC-3	06/15/2022	9045D
06/20/2022	Solid	N3422-02	B Corrosivity	Cool 4 deg C	ROUX02	D11	WC-14A-1A-12	06/20/2022	9045D
06/20/2022	Solid	N3422-08	B Corrosivity	Cool 4 deg C	ROUX02	D11	WC-14A-1A-13	06/20/2022	9045D
06/21/2022	Solid	N3431-02	B Corrosivity	Cool 4 deg C	ROUX02	D11	WC-14B-2-2	06/21/2022	9045D
06/21/2022	Solid	N3431-08	B Corrosivity	Cool 4 deg C	ROUX02	D11	WC-14B-2-3	06/21/2022	9045D
06/21/2022	Solid	N3431-14	B Corrosivity	Cool 4 deg C	ROUX02	D11	WC-14B-2-4	06/21/2022	9045D
06/21/2022	Solid	N3431-20	B Corrosivity	Cool 4 deg C	ROUX02	D11	WC-14B-2-5	06/21/2022	9045D
06/22/2022	Solid	N3447-01	B Corrosivity	Cool 4 deg C	PSEG03	D11	D2356	06/22/2022	9045D
06/22/2022	Water	N3447-02	B Corrosivity	Cool 4 deg C	PSEG03	D11	D2342	06/22/2022	9040C
06/22/2022	Water	N3447-03	B Corrosivity	Cool 4 deg C	PSEG03	D11	D2345	06/22/2022	9040C

Date/Time 6/23/22
 Raw Sample Received by: AS/uc
 Raw Sample Relinquished by: AS (uc)

Date/Time 6/23/22
 Raw Sample Received by: RS (uc)
 Raw Sample Relinquished by: AS/uc