

LB120809


SFAM01.1

INSTRUMENT ID: CV2

Linear

 μ Abs.:

45602



A= 0.0000e+000

B= 2.1958e-004 slope

C= -1.7513e-002 y-intercept

Rho= 0.9999741

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	1/1/1
0.00	0.000	-0.018	-0.018	0	0.000	0					
0.05	0.050					185	0				
0.20	0.200	0.183	-0.017	912	0.0 %	912					
2.50	2.500	2.529	0.029	11599	0.0 %	11599					
5.00	5.000	5.042	0.042	23040	0.0 %	23040					
7.50	7.500	7.468	-0.032	34091	0.0 %	34091					
10.0	10.000	9.996	-0.004	45602	0.0 %	45602					

LB120809
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
0		0	-		0 SFAM01.1	PPB	6/23/2022 16:08	Std
0.2		912	-		0.2 SFAM01.1	PPB	6/23/2022 16:10	Std
2.5		11599	-		2.5 SFAM01.1	PPB	6/23/2022 16:12	Std
5		23040	-		5 SFAM01.1	PPB	6/23/2022 16:15	Std
7.5		34091	-		7.5 SFAM01.1	PPB	6/23/2022 16:17	Std
10		45602	-		10 SFAM01.1	PPB	6/23/2022 16:19	Std
ICV008	ICV008	18163	3.9707 -		SFAM01.1	PPB	6/23/2022 16:22	SMPL
ICB008	ICB008	-59	-0.0305 -		SFAM01.1	PPB	6/23/2022 16:24	SMPL
CCV029	CCV029	22207	4.8587 -		SFAM01.1	PPB	6/23/2022 16:26	SMPL
CCB029	CCB029	-61	-0.0309 -		SFAM01.1	PPB	6/23/2022 16:29	SMPL
PB145616BL	PBS616	-41	-0.0265 -		SFAM01.1	PPB	6/23/2022 16:31	SMPL
N3229-01	GBGY3	3204	0.686 -		SFAM01.1	PPB	6/23/2022 16:33	SMPL
N3229-02	GBGY4	710	0.1384 -		SFAM01.1	PPB	6/23/2022 16:35	SMPL
N3229-03	GBGY4D	788	0.1555 -		SFAM01.1	PPB	6/23/2022 16:38	SMPL
N3229-04	GBGY4S	12150	2.6504 -		SFAM01.1	PPB	6/23/2022 16:40	SMPL
N3229-05	GBGY5	13655	2.9808 -		SFAM01.1	PPB	6/23/2022 16:42	SMPL
PB145617BL	PBS617	-98	-0.039 -		SFAM01.1	PPB	6/23/2022 16:47	SMPL
N3231-01	GBGW8	272	0.0422 -		SFAM01.1	PPB	6/23/2022 16:49	SMPL
N3231-02	GBGW9	187	0.0235 -		SFAM01.1	PPB	6/23/2022 16:51	SMPL
N3231-03	GBGX0	859	0.1711 -		SFAM01.1	PPB	6/23/2022 16:54	SMPL
N3231-04	GBGX0D	844	0.1678 -		SFAM01.1	PPB	6/23/2022 16:56	SMPL
N3231-05	GBGX0S	11895	2.5944 -		SFAM01.1	PPB	6/23/2022 16:58	SMPL
N3231-06	GBGX1	411	0.0727 -		SFAM01.1	PPB	6/23/2022 17:00	SMPL
N3231-07	GBGX2	1043	0.2115 -		SFAM01.1	PPB	6/23/2022 17:03	SMPL
N3231-08	GBGX3	653	0.1259 -		SFAM01.1	PPB	6/23/2022 17:05	SMPL
CCV030	CCV030	22066	4.8277 -		SFAM01.1	PPB	6/23/2022 17:09	SMPL
CCB030	CCB030	-64	-0.0316 -		SFAM01.1	PPB	6/23/2022 17:12	SMPL
N3229-11	GBGY3	19084	4.1729 -		SFAM01.1	PPB	6/23/2022 17:14	SMPL
N3231-14	GBGY0	21542	4.7126 -		SFAM01.1	PPB	6/23/2022 17:21	SMPL
CCV031	CCV031	22706	4.9682 -		SFAM01.1	PPB	6/23/2022 17:27	SMPL
CCB031	CCB031	-70	-0.0329 -		SFAM01.1	PPB	6/23/2022 17:30	SMPL