


LB120817

245.1

INSTRUMENT 11: CV2



Linear

A= 0.0000e+000

B= 2.6239e-004

C= 8.5154e-004

Rho= 0.9998898

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.000	0.000	-3	0.000	-3					
0.05	0.050	0.055	0.005	207	0.0 %	207	0				10
0.2	0.200	0.179	-0.021	679	0.0 %	679					-11
2.5	2.500	2.535	0.035	9659	0.0 %	9659					1
5.0	5.000	5.042	0.042	19212	0.0 %	19212					1
7.5	7.500	7.380	-0.120	28123	0.0 %	28123					-2
10.0	10.000	10.058	0.058	38330	0.0 %	38330					1

LB120817 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 24 Jun 2022 11:19:14

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	59	-3	-	0.0000	245.1	PPB	24 Jun 2022 11:23:36	S	Std
0.05 - 1	50/2	207	-	0.0500	245.1	PPB	24 Jun 2022 11:25:54	S	Std
0.2 - 1	52.5	679	-	0.2000	245.1	PPB	24 Jun 2022 11:28:10	S	Std
2.5 - 1	52.5	9659	-	2.5000	245.1	PPB	24 Jun 2022 11:30:27	S	Std
5.0 - 1	55	19212	-	5.0000	245.1	PPB	24 Jun 2022 11:32:44	S	Std
7.5 - 1	57.5	28123	-	7.5000	245.1	PPB	24 Jun 2022 11:35:00	S	Std
10.0 - 1	510	38330	-	10.0000	245.1	PPB	24 Jun 2022 11:39:54	S	Std
ICV09 - 1	ICV09	14898	3.9100	-245.1		PPB	24 Jun 2022 11:42:39	U	SMPL
ICB09 - 1	ICB09	-68	-0.0170	-245.1		PPB	24 Jun 2022 11:44:53	U	SMPL
CCV32 - 1	CCV32	19289	5.0621	-245.1		PPB	24 Jun 2022 11:47:08	U	SMPL
CCB32 - 1	CCB32	-84	-0.0212	-245.1		PPB	24 Jun 2022 11:49:24	U	SMPL
CRA - 1	CRA	194	0.0518	-245.1		PPB	24 Jun 2022 11:51:39	U	SMPL
HighStd - 1	HighStd	38669	10.1473	-245.1		PPB	24 Jun 2022 11:53:54	U	SMPL
ChkStd - 1	ChkStd	27194	7.1364	-245.1		PPB	24 Jun 2022 11:56:09	U	SMPL
PB145817BL - 1	PBW	-115	-0.0293	-245.1		PPB	24 Jun 2022 11:58:26	U	SMPL
PB145817BS - 1	LCSW	16020	4.2044	-245.1		PPB	24 Jun 2022 12:00:43	U	SMPL
N3297-02 - 1	TI-633	-65	-0.0162	-245.1		PPB	24 Jun 2022 12:02:59	U	SMPL
N3297-02DUP - 1	TI-633DUP	-27	-0.0062	-245.1		PPB	24 Jun 2022 12:05:15	U	SMPL
N3297-02MS - 1	TI-633MS	14101	3.7009	-245.1		PPB	24 Jun 2022 12:07:31	U	SMPL
N3297-02MSD - 1	TI-633MSD	14157	3.7155	-245.1		PPB	24 Jun 2022 12:09:50	U	SMPL
CCV33 - 1	CCV33	19113	5.0160	-245.1		PPB	24 Jun 2022 12:12:06	U	SMPL
CCB33 - 1	CCB33	-80	-0.0201	-245.1		PPB	24 Jun 2022 12:14:21	U	SMPL
N3297-02LX5 - 1		82	0.0224	-245.1		PPB	24 Jun 2022 12:16:37	U	SMPL
N3297-02A - 1		15852	4.1603	-245.1		PPB	24 Jun 2022 12:18:53	U	SMPL
CCV34 - 1	CCV34	19265	5.0559	-245.1		PPB	24 Jun 2022 12:21:09	U	SMPL
CCB34 - 1	CCB34	0	0.0009	-245.1		PPB	24 Jun 2022 12:23:25	U	SMPL