



Analytical Summary Report

Analysis Method: 180.1

ANALYST: Iwona

Parameter: Turbidity

SUPERVISOR REVIEW BY: mohammad

Run Number: LB120824

Reagent/Standard	Lot/Log #
TURBIDITY LOD STD, 0.5NTU	WP97600
Turbidity Calibration std, 0NTU	WP97599
Turbidity Calibration std, 20NTU	WP97594
Turbidity Calibration std, 40NTU	WP97593
Turbidity Calibration std, 1NTU	WP97597
Turbidity Calibration std, 5NTU	WP97596
Turbidity LOQ std, 1.0NTU	WP97601
Turbidity Calibration - CCV std, 10 NTU	WP97595

Intercept: 0.0527

Slope: 0.9301

Regression: 0.9999

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (mg/l)	%D	Anal Date	Anal Time
1	CAL1	0	1	0.144	0.10		06/24/2022	10:20
2	CAL2	1	1	1.150	1.18	-18	06/24/2022	10:24
3	CAL3	5	1	4.370	4.64	7.2	06/24/2022	10:27
4	CAL4	10	1	9.440	10.09	-0.9	06/24/2022	10:30
5	CAL5	20	1	18.600	19.94	0.3	06/24/2022	10:34
6	CAL6	40	1	37.300	40.05	-0.1	06/24/2022	10:38



Analytical Summary Report

Analysis Method: 180.1

ANALYST: Iwona

Parameter: Turbidity

SUPERVISOR REVIEW BY: mohammad

Run Number: LB120824

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	9.310	9.953	06/24/2022	10:42
2	ICB		1	0.152	0.107	06/24/2022	10:45
3	CCV1	10	1	9.330	9.975	06/24/2022	10:48
4	CCB1		1	0.146	0.100	06/24/2022	10:52
5	LB120824BL		1	0.152	0.107	06/24/2022	10:55
6	N3214-07		1	0.524	0.507	06/24/2022	10:58
7	N3214-08		1	1.140	1.169	06/24/2022	11:02
8	CCV2	10	1	9.140	9.770	06/24/2022	11:05
9	CCB2		1	0.144	0.098	06/24/2022	11:08