

LB120839


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

44960



A= 0.0000e+000

B= 2.2431e-004

C= 4.3975e-002

Rho= 0.9995109

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	1/01)
0.0	0.000	0.074	0.074	136	0.000	136					-
0.2	0.200	0.197	-0.003	683	0.0 %	683					-2
2.5	2.500	2.581	0.081	11311	0.0 %	11311					3
5.0	5.000	4.783	-0.217	21129	0.0 %	21129					-4
7.5	7.500	7.435	-0.065	32950	0.0 %	32950					-1
10.0	10.000	10.129	0.129	44960	0.0 %	44960					1

LB120835

LB120835 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 27 Jun 2022 08:58:44					
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	136	-	0.0000	7470A	PPB	27 Jun 2022 09:05:05	S	Std
0.2 - 1	50.2	683	-	0.2000	7470A	PPB	27 Jun 2022 09:07:23	S	Std
2.5 - 1	52.5	11311	-	2.5000	7470A	PPB	27 Jun 2022 09:09:39	S	Std
5.0 - 1	55	21129	-	5.0000	7470A	PPB	27 Jun 2022 09:11:56	S	Std
7.5 - 1	57.5	32950	-	7.5000	7470A	PPB	27 Jun 2022 09:19:19	S	Std
10.0 - 1	510	44960	-	10.0000	7470A	PPB	27 Jun 2022 09:24:09	S	Std
ICV11 - 1	ICV11	15874	3.6046	-	7470A	PPB	27 Jun 2022 09:26:55	U	SMPL
ICB11 - 1	ICB11	-33	0.0366	-	7470A	PPB	27 Jun 2022 09:29:10	U	SMPL
CCV41 - 1	CCV41	20251	4.5864	-	7470A	PPB	27 Jun 2022 09:31:25	U	SMPL
CCB41 - 1	CCB41	-56	0.0314	-	7470A	PPB	27 Jun 2022 09:33:40	U	SMPL
CRA - 1	CRA	660	0.1920	-	7470A	PPB	27 Jun 2022 09:35:54	U	SMPL
HighStd - 1	HighStd	44428	10.0096	-	7470A	PPB	27 Jun 2022 09:38:09	U	SMPL
ChkStd - 1	ChkStd	28195	6.3684	-	7470A	PPB	27 Jun 2022 09:40:24	U	SMPL
PB145833BL - 1	PBW	-99	0.0218	-	7470A	PPB	27 Jun 2022 09:42:41	U	SMPL
PB145833BS - 1	LCSW	15444	3.5082	-	7470A	PPB	27 Jun 2022 09:44:58	U	SMPL
N3454-06 - 1	WC-15-1-2	-37	0.0357	-	7470A	PPB	27 Jun 2022 09:47:14	U	SMPL
N3454-06DUP - 1	WC-15-1-2DUP	-16	0.0404	-	7470A	PPB	27 Jun 2022 09:49:30	U	SMPL
N3454-06MS - 1	WC-15-1-2MS	15037	3.4169	-	7470A	PPB	27 Jun 2022 09:51:47	U	SMPL
N3454-06MSD - 1	WC-15-1-2MSD	15180	3.4490	-	7470A	PPB	27 Jun 2022 09:54:04	U	SMPL
N3454-09 - 1	WC-1A-1A-1	-50	0.0328	-	7470A	PPB	27 Jun 2022 09:56:20	U	SMPL
CCV42 - 1	CCV42	20224	4.5804	-	7470A	PPB	27 Jun 2022 09:58:35	U	SMPL
CCB42 - 1	CCB42	-72	0.0278	-	7470A	PPB	27 Jun 2022 10:00:51	U	SMPL
N3455-04 - 1	SP-EXC-2-1	435	0.1415	-	7470A	PPB	27 Jun 2022 10:03:06	U	SMPL
N3455-10 - 1	SP-EXC-2-2	2004	0.4935	-	7470A	PPB	27 Jun 2022 10:05:22	U	SMPL
N3455-16 - 1	SP-EXC-2-3	346	0.1216	-	7470A	PPB	27 Jun 2022 10:07:37	U	SMPL
N3478-03 - 1	RO-364	3	0.0446	-	7470A	PPB	27 Jun 2022 10:09:53	U	SMPL
PB145801TB - 1	PB145801TB	5	0.0451	-	7470A	PPB	27 Jun 2022 10:12:09	U	SMPL
N3454-06LX5 - 1		-73	0.0276	-	7470A	PPB	27 Jun 2022 10:14:25	U	SMPL
N3454-06A - 1		15840	3.5970	-	7470A	PPB	27 Jun 2022 10:16:42	U	SMPL
CCV43 - 1	CCV43	20376	4.6145	-	7470A	PPB	27 Jun 2022 10:18:59	U	SMPL
CCB43 - 1	CCB43	-75	0.0272	-	7470A	PPB	27 Jun 2022 10:21:16	U	SMPL