



Analytical Summary Report

Analysis Method: 9041A

Analyst By : Guerda

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB120853

Slope : 98.1

BalanceID: WC SC-4

pH Meter ID : WC pH meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2905
BUFFER PH 7.00 GREEN 1PINT PK6	W2906
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.5	4.01	06/27/2022	11:00
2	CAL2	1	Water	NA	NA	21.5	7.00	06/27/2022	11:02
3	CAL3	1	Water	NA	NA	21.5	10.05	06/27/2022	11:04
4	ICV	1	Water	NA	NA	21.8	7.00	06/27/2022	11:06
5	CCV1	1	Water	NA	NA	21.5	2.01	06/27/2022	11:08
6	N3468-01	1	Solid	20	20	19.3	8.10	06/27/2022	11:22
7	N3468-01DUP	1	Solid	20	20	19.3	8.11	06/27/2022	11:22
8	N3469-01	1	Solid	20	20	20.3	8.35	06/27/2022	11:24
9	CCV2	1	Water	NA	NA	20.3	12.03	06/27/2022	11:30

WORKLIST(Hardcopy Internal Chain)

LB126853

Worklist Name : ph-062322

Worklist ID : 160797

Department : Wet-Chemistry

Date : 06-23-2022 08:29:09

Due Date	Matrix	Sample	Test	Raw Sample		Collect Date	Method
				Preservative	Customer Storage Location		
06/22/2022	Solid	N3468-01	pH	Cool 4 deg C	EART12 D11	XRDS-SCREENFILL-CLEAN-FILL-SP-A-COMP	06/22/2022 9041A
06/22/2022	Solid	N3469-01	pH	Cool 4 deg C	EART12 D11	XRDS-SCREENED-CLEAN-FILL-SP-1-COMP	06/22/2022 9041A

Date/Time

6/27/22

Raw Sample Received by:

Shelley

Raw Sample Relinquished by:

KS(wc)

Date/Time

6/27/22

Raw Sample Received by:

Shelley

Raw Sample Relinquished by:

KS(wc)