



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB120855

Analyst By : Guerda

Supervisor Review By : Iwona

Slope : 98.1

pH Meter ID : WC pH meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2905
BUFFER PH 7.00 GREEN 1PINT PK6	W2906
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	24.3	4.01	06/27/2022	08:28
2	CAL2	1	Water	NA	NA	21.5	7.00	06/27/2022	08:30
3	CAL3	1	Water	NA	NA	21.5	10.05	06/27/2022	08:31
4	ICV	1	Water	NA	NA	21.5	6.99	06/27/2022	08:35
5	CCV1	1	Water	NA	NA	21.5	2.01	06/27/2022	08:37
6	N3510-01	1	Water	NA	NA	24.3	5.00	06/27/2022	11:01
7	N3510-01DUP	1	Water	NA	NA	24.3	5.01	06/27/2022	11:05
8	CCV2	1	Water	NA	NA	21.5	12.03	06/27/2022	11:09

WORKLIST(Hardcopy Internal Chain)

L-B120855

WorkList Name : ph-062722 WorkList ID : 160889 Department : Wet-Chemistry Date : 06-27-2022 08:22:25

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
06/22/2022	Water	N3510-01	pH	Cool 4 deg C	STER14	N11	JS10000	06/22/2022	9040C

Date/Time 6/27/22 830
 Raw Sample Received by: SH CWC
 Raw Sample Relinquished by: SH CWC

Date/Time 6/27/22 830
 Raw Sample Received by: SH CWC
 Raw Sample Relinquished by: SH CWC