



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB120856

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2743

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	06/27/2022	10:55
2	N3438-01		1	15	100.00	>212.0	>212.0	06/27/2022	11:20
3	N3438-01DUP		1	15	100.00	>212.0	>212.0	06/27/2022	11:45

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

WorkList Name : fp62722 WorkList ID : 160909 Department : Wet-Chemistry Date : 06-27-2022 10:19:38

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
06/27/2022	Water	N3438-01	Flash Point	Cool 4 deg C	PSEG03	D11	FRAC-TANK-N47935	06/22/2022	1010B

Date/Time 06/27/22 10:25
 Raw Sample Received by: KS(WC)
 Raw Sample Relinquished by: KS(WC)

Date/Time 06/27/22 11:50
 Raw Sample Received by: KS(WC)
 Raw Sample Relinquished by: KS(WC)

LB120456.