


LB120864

7470A

INSTRUMENT ID: CV1

 μ Abs.:

37654



A= 0.0000e+000

B= 2.6552e-004

C= 3.0221e-003

Rho= 0.9999575

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	o/o
0.0	0.000	0.017	0.017	52	0.000	52					0
0.2	0.200	0.200	-0.000	740	0.0 %	740					0
2.5	2.500	2.512	0.012	9450	0.0 %	9450					0
5.0	5.000	4.930	-0.070	18557	0.0 %	18557					-1
7.5	7.500	7.541	0.041	28388	0.0 %	28388					1
10.0	10.000	10.001	0.001	37654	0.0 %	37654					0

LB120864 INSTRUMENT ID : CV1

Method: 7470A	Operator: Admin	Date of Analysis: 28 Jun 2022 10:29:40					Date	Type	Type
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units			
0.0 - 1	50	52	-	0.0000	7470A	PPB	28 Jun 2022 10:42:38	S	Std
0.2 - 1	50.2	740	-	0.2000	7470A	PPB	28 Jun 2022 10:44:55	S	Std
2.5 - 1	52.5	9450	-	2.5000	7470A	PPB	28 Jun 2022 10:47:12	S	Std
5.0 - 1	55	18557	-	5.0000	7470A	PPB	28 Jun 2022 10:52:00	S	Std
7.5 - 1	57.5	28388	-	7.5000	7470A	PPB	28 Jun 2022 10:56:59	S	Std
10.0 - 1	51.8	37654	-	10.0000	7470A	PPB	28 Jun 2022 11:01:47	S	Std
ICV13 - 1	ICV13	14988	3.9826	-	7470A	PPB	28 Jun 2022 11:04:29	U	SMPL
ICB13 - 1	ICB13	-71	-0.0158	-	7470A	PPB	28 Jun 2022 11:06:44	U	SMPL
CCV48 - 1	CCV48	18302	4.8625	-	7470A	PPB	28 Jun 2022 11:09:00	U	SMPL
CCB48 - 1	CCB48	-73	-0.0164	-	7470A	PPB	28 Jun 2022 11:11:14	U	SMPL
CRA - 1	CRA	670	0.1809	-	7470A	PPB	28 Jun 2022 11:13:30	U	SMPL
HighStd - 1	HighStd	37311	9.9097	-	7470A	PPB	28 Jun 2022 11:15:45	U	SMPL
ChkStd - 1	ChkStd	26009	6.9088	-	7470A	PPB	28 Jun 2022 11:18:01	U	SMPL
PB145879BL - 1	PBW	-70	-0.0156	-	7470A	PPB	28 Jun 2022 11:20:18	U	SMPL
PB145879BS - 1	LCSW	14963	3.9759	-	7470A	PPB	28 Jun 2022 11:22:34	U	SMPL
N3472-01 - 1	D2562	-32	-0.0055	-	7470A	PPB	28 Jun 2022 11:24:50	U	SMPL
N3472-02 - 1	D2551	-3	0.0022	-	7470A	PPB	28 Jun 2022 11:27:08	U	SMPL
N3472-03 - 1	D2560	-25	-0.0036	-	7470A	PPB	28 Jun 2022 11:29:25	U	SMPL
N3494-02 - 1	SP-EXC-3	51	0.0166	-	7470A	PPB	28 Jun 2022 11:31:42	U	SMPL
N3494-07 - 1	SP-EXC-1-4	24	0.0094	-	7470A	PPB	28 Jun 2022 11:33:58	U	SMPL
CCV49 - 1	CCV49	18347	4.8745	-	7470A	PPB	28 Jun 2022 11:36:14	U	SMPL
CCB49 - 1	CCB49	-71	-0.0158	-	7470A	PPB	28 Jun 2022 11:38:29	U	SMPL
N3494-12 - 1	SP-EXC-1-5	132	0.0381	-	7470A	PPB	28 Jun 2022 11:40:44	U	SMPL
N3494-17 - 1	SP-EXC-1-6	38	0.0131	-	7470A	PPB	28 Jun 2022 11:43:00	U	SMPL
N3494-22 - 1	SP-EXC-1-3	58	0.0184	-	7470A	PPB	28 Jun 2022 11:45:15	U	SMPL
N3494-27 - 1	SP-EXC-1-2	13	0.0065	-	7470A	PPB	28 Jun 2022 11:47:31	U	SMPL
N3494-32 - 1	SP-EXC-1-1	0	0.0030	-	7470A	PPB	28 Jun 2022 11:49:44	U	SMPL
N3494-32 - 1	SP-EXC-1-1	268	0.0742	-	7470A	PPB	28 Jun 2022 11:59:26	U	SMPL
N3494-32DUP - 1	SP-EXC-1-1DUP	-8	0.0009	-	7470A	PPB	28 Jun 2022 12:01:42	U	SMPL
N3494-32MS - 1	SP-EXC-1-1MS	16087	4.2744	-	7470A	PPB	28 Jun 2022 12:06:32	U	SMPL
N3494-32MSD - 1	SP-EXC-1-1MSD	16071	4.2701	-	7470A	PPB	28 Jun 2022 12:08:48	U	SMPL
CCV50 - 1	CCV50	18183	4.8309	-	7470A	PPB	28 Jun 2022 12:11:04	U	SMPL
CCB50 - 1	CCB50	124	0.0359	-	7470A	PPB	28 Jun 2022 12:34:05	U	SMPL
N3494-32LX5 - 1		52	0.0168	-	7470A	PPB	28 Jun 2022 12:36:21	U	SMPL
N3494-32A - 1		15770	4.1902	-	7470A	PPB	28 Jun 2022 12:38:36	U	SMPL
PB145840TB - 1	PB145840TB	0	0.0030	-	7470A	PPB	28 Jun 2022 12:47:12	U	SMPL
CCV51 - 1	CCV51	18773	4.9876	-	7470A	PPB	28 Jun 2022 12:53:30	U	SMPL
CCB51 - 1	CCB51	107	0.0314	-	7470A	PPB	28 Jun 2022 13:09:19	U	SMPL