



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB120865

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2743

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.1	06/28/2022	09:55
2	N3471-15		1	12	100.00	>212.0	>212.0	06/28/2022	10:20
3	N3471-15DUP		1	12	100.00	>212.0	>212.0	06/28/2022	10:45
4	N3471-16		1	15	100.00	>212.0	>212.0	06/28/2022	11:10
5	N3471-17		1	17	100.00	>212.0	>212.0	06/28/2022	11:35
6	N3480-01		1	18	100.00	>212.0	>212.0	06/28/2022	11:50
7	N3480-02		1	16	100.00	>212.0	>212.0	06/28/2022	12:15
8	N3480-03		1	20	100.00	>212.0	>212.0	06/28/2022	12:40

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB120865

WorkList Name : fp62822 WorkList ID : 160926 Department : Wet-Chemistry Date : 06-28-2022 09:27:29

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample		Collect Date	Method
						Storage Location	Customer Sample		
06/28/2022	Water	N3471-15	Flash Point	Cool 4 deg C	PSEG03	F11	COMP1-WATER	06/23/2022	1010B
06/28/2022	Water	N3471-16	Flash Point	Cool 4 deg C	PSEG03	F11	COMP2-WATER	06/23/2022	1010B
06/28/2022	Water	N3471-17	Flash Point	Cool 4 deg C	PSEG03	F11	COMP3-WATER	06/23/2022	1010B
06/28/2022	Water	N3480-01	Flash Point	Cool 4 deg C	PSEG03	F11	40814	06/23/2022	1010B
06/28/2022	Water	N3480-02	Flash Point	Cool 4 deg C	PSEG03	F11	40815	06/23/2022	1010B
06/28/2022	Water	N3480-03	Flash Point	Cool 4 deg C	PSEG03	F11	31522B	06/23/2022	1010B

Date/Time 06/28/22 09:30
 Raw Sample Received by: KS (WC)
 Raw Sample Relinquished by: [Signature]

Date/Time 06/28/22 12:45
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: KS (WC)