

LB120876


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

53053



A= 0.0000e+000

B= 1.9081e-004

C= -2.2495e-001

Rho= 0.9998346

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.027	-0.027	1036	0.000	1036					-
0.2	0.200	0.168	-0.032	2057	0.0 %	2057					-16
2.5	2.500	2.539	0.039	14484	0.0 %	14484					2
5.0	5.000	5.008	0.008	27425	0.0 %	27425					0
7.5	7.500	7.615	0.115	41089	0.0 %	41089					2
10.0	10.000	9.898	-0.102	53053	0.0 %	53053					-1

LB120876 INSTRUMENT ID : CV1

Method: 7471B		Operator: Admin		Date of Analysis: 29 Jun 2022 06:59:37				
Sample ID	Extended ID	μ Abs.	Conc.	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	50	1036	-	0.00007471B	PPB	29 Jun 2022 07:14:31	S	Std
0.2 - 1	50.2	2057	-	0.20007471B	PPB	29 Jun 2022 07:16:48	S	Std
2.5 - 1	52.5	14484	-	2.50007471B	PPB	29 Jun 2022 07:19:05	S	Std
5.0 - 1	55	27425	-	5.00007471B	PPB	29 Jun 2022 07:21:22	S	Std
7.5 - 1	57.5	41089	-	7.50007471B	PPB	29 Jun 2022 07:23:39	S	Std
10.0 - 1	510	53053	-	10.00007471B	PPB	29 Jun 2022 07:28:32	S	Std
ICV17 - 1	ICV17	22728	4.1117	-7471B	PPB	29 Jun 2022 07:31:21	U	SMPL
ICB17 - 1	ICB17	818	-0.0689	-7471B	PPB	29 Jun 2022 07:36:18	U	SMPL
CCV56 - 1	CCV56	26838	4.8959	-7471B	PPB	29 Jun 2022 07:38:33	U	SMPL
CCB56 - 1	CCB56	893	-0.0546	-7471B	PPB	29 Jun 2022 07:40:48	U	SMPL
CRA - 1	CRA	1934	0.1441	-7471B	PPB	29 Jun 2022 07:48:16	U	SMPL
HighStd - 1	HighStd	52873	9.8636	-7471B	PPB	29 Jun 2022 07:50:30	U	SMPL
ChkStd - 1	ChkStd	37964	7.0189	-7471B	PPB	29 Jun 2022 07:52:46	U	SMPL
PB145905BL - 1	PBS	1012	-0.0319	-7471B	PPB	29 Jun 2022 07:55:03	U	SMPL
PB145905BS - 1	LCSS	20727	3.7299	-7471B	PPB	29 Jun 2022 07:57:21	U	SMPL
N3492-02 - 1	SOIL-PILE	1457	0.0531	-7471B	PPB	29 Jun 2022 07:59:37	U	SMPL
N3494-01 - 1	SP-EXC-3	91123	17.1620	-7471B	PPB	29 Jun 2022 08:01:54	U	SMPL
N3494-06 - 1	SP-EXC-1-4	127221	24.0497	-7471B	PPB	29 Jun 2022 08:04:11	U	SMPL
N3494-11 - 1	SP-EXC-1-5	119407	22.5588	-7471B	PPB	29 Jun 2022 08:06:31	U	SMPL
N3494-16 - 1	SP-EXC-1-6	142043	26.8779	-7471B	PPB	29 Jun 2022 08:08:50	U	SMPL
CCV57 - 1	CCV57	26511	4.8335	-7471B	PPB	29 Jun 2022 08:11:08	U	SMPL
CCB57 - 1	CCB57	780	-0.0761	-7471B	PPB	29 Jun 2022 08:13:27	U	SMPL
N3494-21 - 1	SP-EXC-1-3	100824	19.0130	-7471B	PPB	29 Jun 2022 08:15:44	U	SMPL
N3494-26 - 1	SP-EXC-1-2	114355	21.5948	-7471B	PPB	29 Jun 2022 08:17:59	U	SMPL
N3494-31 - 1	SP-EXC-1-1	135365	25.6037	-7471B	PPB	29 Jun 2022 08:20:18	U	SMPL
N3499-02 - 1	31547	1518	0.0647	-7471B	PPB	29 Jun 2022 08:22:37	U	SMPL
N3499-03 - 1	BUR-1206	1467	0.0550	-7471B	PPB	29 Jun 2022 08:24:57	U	SMPL
N3499-03DUP - 1	BUR-1206DUP	1299	0.0229	-7471B	PPB	29 Jun 2022 08:27:13	U	SMPL
N3499-03MS - 1	BUR-1206MS	23987	4.3519	-7471B	PPB	29 Jun 2022 08:29:30	U	SMPL
N3499-03MSD - 1	BUR-1206MSD	21940	3.9614	-7471B	PPB	29 Jun 2022 08:31:47	U	SMPL
N3502-01 - 1	D-46-UC	3979	0.5343	-7471B	PPB	29 Jun 2022 08:34:05	U	SMPL
N3502-02 - 1	D-46-LC	3486	0.4402	-7471B	PPB	29 Jun 2022 08:36:22	U	SMPL
CCV58 - 1	CCV58	26635	4.8572	-7471B	PPB	29 Jun 2022 08:38:37	U	SMPL
CCB58 - 1	CCB58	578	-0.1147	-7471B	PPB	29 Jun 2022 08:40:53	U	SMPL
N3502-03 - 1	D-42-UC	2305	0.2149	-7471B	PPB	29 Jun 2022 08:43:10	U	SMPL
N3502-04 - 1	D-42-LC	2736	0.2971	-7471B	PPB	29 Jun 2022 08:45:25	U	SMPL
N3502-05 - 1	D-43-UC	3484	0.4398	-7471B	PPB	29 Jun 2022 08:47:41	U	SMPL
N3502-06 - 1	D-43-LC	4455	0.6251	-7471B	PPB	29 Jun 2022 08:49:56	U	SMPL
N3502-07 - 1	D-44-UC	4174	0.5715	-7471B	PPB	29 Jun 2022 08:52:12	U	SMPL
N3502-08 - 1	D-44-LC	4918	0.7134	-7471B	PPB	29 Jun 2022 08:54:29	U	SMPL
N3503-01 - 1	OR-03-062422	2243	0.2030	-7471B	PPB	29 Jun 2022 08:56:45	U	SMPL
N3499-03LX5 - 1		267	-0.1740	-7471B	PPB	29 Jun 2022 08:59:02	U	SMPL
N3499-03A - 1		22777	4.1211	-7471B	PPB	29 Jun 2022 09:01:19	U	SMPL
CCV59 - 1	CCV59	26559	4.8427	-7471B	PPB	29 Jun 2022 09:03:35	U	SMPL
CCB59 - 1	CCB59	800	-0.0723	-7471B	PPB	29 Jun 2022 09:05:52	U	SMPL
N3494-01DLX5 - 1		19162	3.4313	-7471B	PPB	29 Jun 2022 09:08:08	U	SMPL
N3494-06DLX5 - 1		26940	4.9154	-7471B	PPB	29 Jun 2022 09:10:24	U	SMPL
N3494-11DLX5 - 1		24939	4.5336	-7471B	PPB	29 Jun 2022 09:12:41	U	SMPL
N3494-16DLX5 - 1		27299	4.9839	-7471B	PPB	29 Jun 2022 09:14:57	U	SMPL
N3494-21DLX5 - 1		21637	3.9035	-7471B	PPB	29 Jun 2022 09:17:13	U	SMPL
N3494-26DLX5 - 1		23885	4.3325	-7471B	PPB	29 Jun 2022 09:19:31	U	SMPL
N3494-31DLX5 - 1		28552	5.2230	-7471B	PPB	29 Jun 2022 09:21:48	U	SMPL
CCV60 - 1	CCV60	26539	4.8389	-7471B	PPB	29 Jun 2022 09:24:06	U	SMPL
CCB60 - 1	CCB60	820	-0.0685	-7471B	PPB	29 Jun 2022 09:26:24	U	SMPL