



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Guerda

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB120879

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2743

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	28.00	82.4	82.1	06/29/2022	09:40
2	N3436-01		1	15	100.00	>212.0	>212.0	06/29/2022	10:10
3	N3436-01DUP		1	15	100.00	>212.0	>212.0	06/29/2022	10:35
4	N3447-02		1	17	100.00	>212.0	>212.0	06/29/2022	11:05
5	N3447-03		1	18	100.00	>212.0	>212.0	06/29/2022	11:32
6	N3491-01		1	16	100.00	>212.0	>212.0	06/29/2022	11:55
7	N3491-02		1	12	100.00	>212.0	>212.0	06/29/2022	12:20
8	N3499-04		1	15	100.00	>212.0	>212.0	06/29/2022	12:45
9	N3504-07		1	18	100.00	>212.0	>212.0	06/29/2022	13:10
10	N3504-08		1	17	100.00	>212.0	>212.0	06/29/2022	13:35

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB 120879

WorkList Name : FP62922 WorkList ID : 160962 Department : Wet-Chemistry Date : 06-28-2022 16:54:26

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
06/29/2022	Water	N3436-01	Flash Point	Cool 4 deg C	PSEG03	D11	31677	06/23/2022	1010B
06/29/2022	Water	N3447-02	Flash Point	Cool 4 deg C	PSEG03	D11	D2342	06/22/2022	1010B
06/29/2022	Water	N3447-03	Flash Point	Cool 4 deg C	PSEG03	D11	D2345	06/22/2022	1010B
06/29/2022	Water	N3491-01	Flash Point	Cool 4 deg C	PSEG03	F11	AUD-1728	06/24/2022	1010B
06/29/2022	Water	N3491-02	Flash Point	Cool 4 deg C	PSEG03	F11	AUD-1613	06/24/2022	1010B
06/29/2022	Water	N3499-04	Flash Point	Cool 4 deg C	PSEG03	F11	31052-WATER	06/24/2022	1010B
06/29/2022	Water	N3504-07	Flash Point	Cool 4 deg C	PSEG03	F11	TRE-1613-15	06/24/2022	1010B
06/29/2022	Water	N3504-08	Flash Point	Cool 4 deg C	PSEG03	F11	TRE-1616-1618	06/24/2022	1010B

Date/Time 06/29/22 09:15
 Raw Sample Received by: SM(wc)
 Raw Sample Relinquished by: KS(wc)

Date/Time 06/29/22 13:40
 Raw Sample Received by: KS(wc)
 Raw Sample Relinquished by: SM(wc)