



Analytical Summary Report

Analysis Method: SM 4500-H B

Parameter: pH

Run Number: LB120894

Analyst By : JIGNESH

Supervisor Review By : mohammad

Slope : 99.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2905
BUFFER PH 7.00 GREEN 1PINT PK6	W2906
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.00	06/30/2022	09:45
2	CAL2	1	Water	NA	NA	20.2	7.01	06/30/2022	09:46
3	CAL3	1	Water	NA	NA	20.2	10.03	06/30/2022	09:47
4	ICV	1	Water	NA	NA	20.3	6.98	06/30/2022	09:50
5	CCV1	1	Water	NA	NA	20.4	2.01	06/30/2022	09:52
6	N3546-01	1	Water	NA	NA	20.1	5.03	06/30/2022	10:00
7	N3546-02	1	Water	NA	NA	20.1	5.28	06/30/2022	10:02
8	N3546-03	1	Water	NA	NA	20.2	5.26	06/30/2022	10:03
9	N3546-04	1	Water	NA	NA	20.2	5.29	06/30/2022	10:05
10	N3546-04DUP	1	Water	NA	NA	20.3	5.30	06/30/2022	10:06
11	CCV2	1	Water	NA	NA	20.3	12.03	06/30/2022	10:10

18120896

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph n3546 WorkList ID : 161023 Department : Wet-Chemistry Date : 06-30-2022 09:40:43

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
06/29/2022	Water	N3546-01	pH	Cool 4 deg C	NYUL01	F11	TH-4-30-DI-1	06/29/2022	SM 4500-H B
06/29/2022	Water	N3546-02	pH	Cool 4 deg C	NYUL01	F11	TH-4-30-DI-2	06/29/2022	SM 4500-H B
06/29/2022	Water	N3546-03	pH	Cool 4 deg C	NYUL01	F11	TH-4-30-DI-3	06/29/2022	SM 4500-H B
06/29/2022	Water	N3546-04	pH	Cool 4 deg C	NYUL01	F11	TH-4-04-DI-4	06/29/2022	SM 4500-H B

Date/Time 06/30/22 09:42
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]

Date/Time 06/30/22 11:30
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]