

LB120896


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

56656



A= 0.0000e+000

B= 1.7750e-004

C= -4.4391e-002

Rho= 0.9999307

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%b
0.0	0.000	-0.019	-0.019	142	0.000	142					1
0.2	0.200	0.206	0.006	1409	0.0 %	1409					3
2.5	2.500	2.489	-0.011	14275	0.0 %	14275					0
5.0	5.000	5.079	0.079	28862	0.0 %	28862					2
7.5	7.500	7.433	-0.067	42126	0.0 %	42126					0
10.0	10.000	10.012	0.012	56656	0.0 %	56656					0

LN120896 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 30 Jun 2022 10:48:16

Sample ID	Extended ID	μ Abs.	Conc.	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	50	142	-	0.00007471B	PPB	30 Jun 2022 10:54:02	S	Std
0.2 - 1	50.2	1409	-	0.20007471B	PPB	30 Jun 2022 10:56:19	S	Std
2.5 - 1	52.5	14275	-	2.50007471B	PPB	30 Jun 2022 10:58:37	S	Std
5.0 - 1	55	28862	-	5.00007471B	PPB	30 Jun 2022 11:00:54	S	Std
7.5 - 1	57.5	42126	-	7.50007471B	PPB	30 Jun 2022 11:03:10	S	Std
10.0 - 1	510	56656	-	10.00007471B	PPB	30 Jun 2022 11:08:02	S	Std
ICV19 - 1	ICV19	23318	4.0946	-7471B	PPB	30 Jun 2022 11:10:39	U	SMPL
ICB19 - 1	ICB19	10	-0.0426	-7471B	PPB	30 Jun 2022 11:12:54	U	SMPL
CCV65 - 1	CCV65	28324	4.9832	-7471B	PPB	30 Jun 2022 11:15:10	U	SMPL
CCB65 - 1	CCB65	46	-0.0362	-7471B	PPB	30 Jun 2022 11:17:25	U	SMPL
CRA - 1	CRA	1241	0.1759	-7471B	PPB	30 Jun 2022 11:19:40	U	SMPL
ChkStd - 1	ChkStd	40425	7.1312	-7471B	PPB	30 Jun 2022 11:24:11	U	SMPL
PB145961BL - 1	PBS	-21	-0.0481	-7471B	PPB	30 Jun 2022 11:26:26	U	SMPL
PB145961BS - 1	LCSS	22810	4.0045	-7471B	PPB	30 Jun 2022 11:28:44	U	SMPL
N3512-01 - 1	OR-3-062722	3150	0.5147	-7471B	PPB	30 Jun 2022 11:31:00	U	SMPL
N3512-01DUP - 1	OR-3-062722DUP	2705	0.4358	-7471B	PPB	30 Jun 2022 11:33:18	U	SMPL
N3512-01MS - 1	OR-3-062722MS	25952	4.5622	-7471B	PPB	30 Jun 2022 11:35:35	U	SMPL
N3512-01MSD - 1	OR-3-062722MSD	24659	4.3327	-7471B	PPB	30 Jun 2022 11:37:52	U	SMPL
N3516-01 - 1	R-1	11011	1.9101	-7471B	PPB	30 Jun 2022 11:40:08	U	SMPL
CCV66 - 1	CCV66	28767	5.0618	-7471B	PPB	30 Jun 2022 11:42:24	U	SMPL
CCB66 - 1	CCB66	12	-0.0423	-7471B	PPB	30 Jun 2022 11:44:40	U	SMPL
N3524-03 - 1	31493	4127	0.6882	-7471B	PPB	30 Jun 2022 11:46:57	U	SMPL
N3524-04 - 1	31546	3324	0.5456	-7471B	PPB	30 Jun 2022 11:49:12	U	SMPL
N3524-05 - 1	31494	2954	0.4800	-7471B	PPB	30 Jun 2022 11:51:28	U	SMPL
N3531-01 - 1	W-1	1789	0.2732	-7471B	PPB	30 Jun 2022 11:53:43	U	SMPL
N3531-02 - 1	W-2	1685	0.2547	-7471B	PPB	30 Jun 2022 11:55:59	U	SMPL
N3539-01 - 1	CL-02-062922	2397	0.3811	-7471B	PPB	30 Jun 2022 11:58:15	U	SMPL
N3547-01 - 1	T-1	1329	0.1915	-7471B	PPB	30 Jun 2022 12:00:32	U	SMPL
N3548-01 - 1	V-1	731	0.0854	-7471B	PPB	30 Jun 2022 12:02:49	U	SMPL
N3549-01 - 1	ML-1	2087	0.3261	-7471B	PPB	30 Jun 2022 12:05:05	U	SMPL
N3549-02 - 1	ML-2	1738	0.2641	-7471B	PPB	30 Jun 2022 12:07:21	U	SMPL
CCV67 - 1	CCV67	28944	5.0933	-7471B	PPB	30 Jun 2022 12:09:36	U	SMPL
CCB67 - 1	CCB67	-39	-0.0513	-7471B	PPB	30 Jun 2022 12:11:52	U	SMPL
N3550-01 - 1	HR-02-06292022	15180	2.6501	-7471B	PPB	30 Jun 2022 12:14:08	U	SMPL
N3550-03 - 1	HR-03-06292022	4137	0.6899	-7471B	PPB	30 Jun 2022 12:16:24	U	SMPL
N3512-01LX5 - 1		689	0.0779	-7471B	PPB	30 Jun 2022 12:18:39	U	SMPL
N3512-01A - 1		25805	4.5361	-7471B	PPB	30 Jun 2022 12:20:55	U	SMPL
HighStd - 1	HighStd	55960	9.8887	-7471B	PPB	30 Jun 2022 12:23:10	U	SMPL
CCV68 - 1	CCV68	28044	4.9335	-7471B	PPB	30 Jun 2022 12:25:28	U	SMPL
CCB68 - 1	CCB68	-41	-0.0517	-7471B	PPB	30 Jun 2022 12:27:46	U	SMPL