

LB120903


SFAM01.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

44864



A= 0.0000e+000

B= 2.2322e-004 slope

C= 1.7103e-002 intercept

Rho= 0.9999805

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.012	0.012	-22	0.000	-22					1
0.05	0.050					185	0				1
0.20	0.200	0.204	0.004	836	0.0 %	836					2
2.50	2.500	2.490	-0.010	11078	0.0 %	11078					0
5.00	5.000	5.006	0.006	22348	0.0 %	22348					0
7.50	7.500	7.457	-0.043	33328	0.0 %	33328					0
10.0	10.000	10.032	0.032	44864	0.0 %	44864					0

LB120903
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	-22	-		0 SFAM01.1	PPB	6/30/2022 14:15	Std
	0.2 S01	836	-		0.2 SFAM01.1	PPB	6/30/2022 14:17	Std
	2.5 S02	11078	-		2.5 SFAM01.1	PPB	6/30/2022 14:19	Std
	5 S03	22348	-		5 SFAM01.1	PPB	6/30/2022 14:22	Std
	7.5 S04	33328	-		7.5 SFAM01.1	PPB	6/30/2022 14:24	Std
	10 S05	44864	-		10 SFAM01.1	PPB	6/30/2022 14:26	Std
ICV020	ICV020	16621	3.7273	-	SFAM01.1	PPB	6/30/2022 14:29	SMPL
ICB020	ICB020	-73	0.0008	-	SFAM01.1	PPB	6/30/2022 14:31	SMPL
CCV069	CCV069	22071	4.9439	-	SFAM01.1	PPB	6/30/2022 14:33	SMPL
CCB069	CCB069	-66	0.0024	-	SFAM01.1	PPB	6/30/2022 14:35	SMPL
PB145962BL	PBW962	-45	0.0071	-	SFAM01.1	PPB	6/30/2022 14:38	SMPL
N3403-09	MC0AD2	-52	0.0055	-	SFAM01.1	PPB	6/30/2022 14:40	SMPL
N3403-10	MC0AD5	-46	0.0068	-	SFAM01.1	PPB	6/30/2022 14:42	SMPL
PB145963BL	PBW963	-14	0.014	-	SFAM01.1	PPB	6/30/2022 14:45	SMPL
N3405-01	MC0AD4	-21	0.0124	-	SFAM01.1	PPB	6/30/2022 14:47	SMPL
N3405-02	MC0AD4D	-19	0.0129	-	SFAM01.1	PPB	6/30/2022 14:49	SMPL
N3405-03	MC0AD4S	4320	0.9814	-	SFAM01.1	PPB	6/30/2022 14:51	SMPL
N3405-04	MC0AC2	-2	0.0167	-	SFAM01.1	PPB	6/30/2022 14:54	SMPL
N3405-05	MC0AC9	6	0.0184	-	SFAM01.1	PPB	6/30/2022 14:56	SMPL
N3405-06	MC0AD0	11	0.0196	-	SFAM01.1	PPB	6/30/2022 14:58	SMPL
N3405-07	MC0AF5	-52	0.0055	-	SFAM01.1	PPB	6/30/2022 15:00	SMPL
N3405-08	MC0AD1	-59	0.0039	-	SFAM01.1	PPB	6/30/2022 15:03	SMPL
N3405-09	MC0AF3	-5	0.016	-	SFAM01.1	PPB	6/30/2022 15:05	SMPL
N3405-10	MC0AD3	-3	0.0164	-	SFAM01.1	PPB	6/30/2022 15:07	SMPL
N3405-11	MC0AF8	-13	0.0142	-	SFAM01.1	PPB	6/30/2022 15:09	SMPL
N3405-12	MC0AE8	-50	0.0059	-	SFAM01.1	PPB	6/30/2022 15:12	SMPL
N3405-13	MC0AE7	116	0.043	-	SFAM01.1	PPB	6/30/2022 15:14	SMPL
N3405-14	MC0AF0	-51	0.0057	-	SFAM01.1	PPB	6/30/2022 15:16	SMPL
N3405-15	MC0AD6	-35	0.0093	-	SFAM01.1	PPB	6/30/2022 15:18	SMPL
N3405-16	MC0AD9	-25	0.0115	-	SFAM01.1	PPB	6/30/2022 15:21	SMPL
CCV070	CCV070	21654	4.8508	-	SFAM01.1	PPB	6/30/2022 15:23	SMPL
CCB070	CCB070	-66	0.0024	-	SFAM01.1	PPB	6/30/2022 15:25	SMPL
N3405-17	MC0AE9	-43	0.0075	-	SFAM01.1	PPB	6/30/2022 15:28	SMPL
N3405-18	MC0AE5	9	0.0191	-	SFAM01.1	PPB	6/30/2022 15:30	SMPL
N3405-19	MC0AA3	-38	0.0086	-	SFAM01.1	PPB	6/30/2022 15:32	SMPL
N3405-20	MC0AD8	-2	0.0167	-	SFAM01.1	PPB	6/30/2022 15:34	SMPL
N3405-21	MC0AF9	-38	0.0086	-	SFAM01.1	PPB	6/30/2022 15:37	SMPL
PB145965BL	PBW965	-23	0.012	-	SFAM01.1	PPB	6/30/2022 15:39	SMPL
N3511-01	MC2B01	-40	0.0082	-	SFAM01.1	PPB	6/30/2022 15:41	SMPL
N3511-02	MC2B02	-52	0.0055	-	SFAM01.1	PPB	6/30/2022 15:43	SMPL
N3511-03	MC2B03	-22	0.0122	-	SFAM01.1	PPB	6/30/2022 15:46	SMPL
N3511-04	MC2B04	-54	0.005	-	SFAM01.1	PPB	6/30/2022 15:48	SMPL
N3511-05	MC2B05	-92	-0.0034	-	SFAM01.1	PPB	6/30/2022 16:00	SMPL
N3511-06	MC2B06	-107	-0.0068	-	SFAM01.1	PPB	6/30/2022 16:02	SMPL
N3511-07	MC2B06S	4548	1.0323	-	SFAM01.1	PPB	6/30/2022 16:04	SMPL
N3511-08	MC2B06D	-76	0.0001	-	SFAM01.1	PPB	6/30/2022 16:06	SMPL

LB120903

INSTRUMENT ID : CV1

CCB071	CCB071	21957	4.9184 -	SFAM01.1	PPB	6/30/2022 16:09 SMPL
CCB071	CCB071	83	0.0356 -	SFAM01.1	PPB	6/30/2022 16:11 SMPL
N3511-09	MC2B07	-39	0.0084 -	SFAM01.1	PPB	6/30/2022 16:13 SMPL
N3511-10	MC2B08	9	0.0191 -	SFAM01.1	PPB	6/30/2022 16:15 SMPL
N3511-11	MC2B09	-11	0.0146 -	SFAM01.1	PPB	6/30/2022 16:18 SMPL
N3511-12	MC2B10	-36	0.0091 -	SFAM01.1	PPB	6/30/2022 16:20 SMPL
N3511-13	MC2B11	-60	0.0037 -	SFAM01.1	PPB	6/30/2022 16:22 SMPL
N3511-14	MC2B12	-47	0.0066 -	SFAM01.1	PPB	6/30/2022 16:25 SMPL
N3511-15	MC2B13	-22	0.0122 -	SFAM01.1	PPB	6/30/2022 16:27 SMPL
CCV072	CCV072	21630	4.8454 -	SFAM01.1	PPB	6/30/2022 16:29 SMPL
CCB072	CCB072	-67	0.0021 -	SFAM01.1	PPB	6/30/2022 16:31 SMPL