

LB120912


SFAM01.1

INSTRUMENT ID: C112

Linear

μ Abs.:

44085



A= 0.0000e+000

B= 2.2895e-004 slope

C= 1.7193e-002 y-intercept

Rho= 0.9998142

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%OD
0.00	0.000	0.040	0.040	100	0.000	100					11
0.05	0.050					185	0				13
0.20	0.200	0.226	0.026	912	0.0 %	912					0
2.50	2.500	2.498	-0.002	10836	0.0 %	10836					11
5.00	5.000	4.930	-0.070	21459	0.0 %	21459					11
7.50	7.500	7.395	-0.105	32225	0.0 %	32225					11
10.0	10.000	10.110	0.110	44085	0.0 %	44085					11

LB120912
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	100	-		0 SFAM01.1	PPB	7/1/2022 10:47	Std
	0.2 S01	912	-		0.2 SFAM01.1	PPB	7/1/2022 10:49	Std
	2.5 S02	10836	-		2.5 SFAM01.1	PPB	7/1/2022 10:52	Std
	5 S03	21459	-		5 SFAM01.1	PPB	7/1/2022 10:54	Std
	7.5 S04	32225	-		7.5 SFAM01.1	PPB	7/1/2022 10:56	Std
	10 S05	44085	-		10 SFAM01.1	PPB	7/1/2022 10:58	Std
ICV021	ICV021	17040	3.9185	-	SFAM01.1	PPB	7/1/2022 11:02	SMPL
ICB021	ICB021	-85	-0.0023	-	SFAM01.1	PPB	7/1/2022 11:04	SMPL
CCV073	CCV073	21735	4.9934	-	SFAM01.1	PPB	7/1/2022 11:06	SMPL
CCB073	CCB073	-70	0.0012	-	SFAM01.1	PPB	7/1/2022 11:09	SMPL
PB145974BL	PBW974	-9	0.0151	-	SFAM01.1	PPB	7/1/2022 11:11	SMPL
N3406-01	MCOAF6	-61	0.0032	-	SFAM01.1	PPB	7/1/2022 11:13	SMPL
N3406-02	MCOAE0	79	0.0353	-	SFAM01.1	PPB	7/1/2022 11:15	SMPL
N3406-03	MCOAE0D	57	0.0302	-	SFAM01.1	PPB	7/1/2022 11:18	SMPL
N3406-04	MCOAE0S	4090	0.9536	-	SFAM01.1	PPB	7/1/2022 11:20	SMPL
N3406-05	MCOAG0	-66	0.0021	-	SFAM01.1	PPB	7/1/2022 11:22	SMPL
N3406-06	MCOAD7	-24	0.0117	-	SFAM01.1	PPB	7/1/2022 11:24	SMPL
N3406-07	MCOAE1	-42	0.0076	-	SFAM01.1	PPB	7/1/2022 11:27	SMPL
N3406-08	MCOAE2	-60	0.0035	-	SFAM01.1	PPB	7/1/2022 11:29	SMPL
N3406-09	MCOAE3	-55	0.0046	-	SFAM01.1	PPB	7/1/2022 11:31	SMPL
N3406-10	MCOAE6	-21	0.0124	-	SFAM01.1	PPB	7/1/2022 11:33	SMPL
N3406-11	MCOAF1	-58	0.0039	-	SFAM01.1	PPB	7/1/2022 11:36	SMPL
N3406-12	MCOAF2	-48	0.0062	-	SFAM01.1	PPB	7/1/2022 11:38	SMPL
N3406-13	MCOAG1	-30	0.0103	-	SFAM01.1	PPB	7/1/2022 11:40	SMPL
N3406-14	MCOAG2	-56	0.0044	-	SFAM01.1	PPB	7/1/2022 11:42	SMPL
PB145975BL	PBW975	5	0.0183	-	SFAM01.1	PPB	7/1/2022 11:45	SMPL
N3461-01	MESQQ1	11	0.0197	-	SFAM01.1	PPB	7/1/2022 11:47	SMPL
N3461-02	MESQQ3	21	0.022	-	SFAM01.1	PPB	7/1/2022 11:49	SMPL
N3461-03	MESQQ5	294	0.0845	-	SFAM01.1	PPB	7/1/2022 11:52	SMPL
N3461-04	MESQQ5D	287	0.0829	-	SFAM01.1	PPB	7/1/2022 11:54	SMPL
CCV074	CCV074	22159	5.0905	-	SFAM01.1	PPB	7/1/2022 11:56	SMPL
CCB074	CCB074	-81	-0.0014	-	SFAM01.1	PPB	7/1/2022 11:58	SMPL
N3461-05	MESQQ5S	4719	1.0976	-	SFAM01.1	PPB	7/1/2022 12:01	SMPL
N3461-06	MESQQ7	657	0.1676	-	SFAM01.1	PPB	7/1/2022 12:03	SMPL
N3461-07	MESQQ9	245	0.0733	-	SFAM01.1	PPB	7/1/2022 12:05	SMPL
N3461-08	MESQR0	141	0.0495	-	SFAM01.1	PPB	7/1/2022 12:07	SMPL
PB145976BL	PBW976	250	0.0744	-	SFAM01.1	PPB	7/1/2022 12:10	SMPL
N3500-01	MC1B21	142	0.0497	-	SFAM01.1	PPB	7/1/2022 12:12	SMPL
N3500-02	MC1B21D	140	0.0492	-	SFAM01.1	PPB	7/1/2022 12:14	SMPL
N3500-03	MC1B21S	4402	1.025	-	SFAM01.1	PPB	7/1/2022 12:16	SMPL
N3500-04	MC1B22	139	0.049	-	SFAM01.1	PPB	7/1/2022 12:19	SMPL
N3500-05	MC1B23	156	0.0529	-	SFAM01.1	PPB	7/1/2022 12:21	SMPL
N3500-06	MC1B24	630	0.1614	-	SFAM01.1	PPB	7/1/2022 12:23	SMPL
N3500-07	MC1B25	280	0.0813	-	SFAM01.1	PPB	7/1/2022 12:26	SMPL
N3500-08	MC1B26	235	0.071	-	SFAM01.1	PPB	7/1/2022 12:28	SMPL
N3500-09	MC1B27	196	0.0621	-	SFAM01.1	PPB	7/1/2022 12:30	SMPL

LB120912

INSTRUMENT ID : CV1

N3500-10	MC1B28	217	0.0669 -	SFAM01.1	PPB	7/1/2022 12:32 SMPL
N3500-11	MC1B29	782	0.1962 -	SFAM01.1	PPB	7/1/2022 12:35 SMPL
PB145977BL	PBW977	308	0.0877 -	SFAM01.1	PPB	7/1/2022 12:37 SMPL
N3356-01	MEY0C3	217	0.0669 -	SFAM01.1	PPB	7/1/2022 12:39 SMPL
N3356-02	MEY0C7	177	0.0577 -	SFAM01.1	PPB	7/1/2022 12:41 SMPL
N3356-03	MEY0C5	182	0.0589 -	SFAM01.1	PPB	7/1/2022 12:44 SMPL
CCV075	CCV075	23050	5.2945 -	SFAM01.1	PPB	7/1/2022 12:54 SMPL
CCB075	CCB075	-4	0.0163 -	SFAM01.1	PPB	7/1/2022 12:56 SMPL
N3356-04	MEY0C6	223	0.0682 -	SFAM01.1	PPB	7/1/2022 12:58 SMPL
N3356-05	MEY0C9	328	0.0923 -	SFAM01.1	PPB	7/1/2022 13:07 SMPL
N3356-06	MEY0C9D	201	0.0632 -	SFAM01.1	PPB	7/1/2022 13:09 SMPL
N3356-07	MEY0C9S	4898	1.1386 -	SFAM01.1	PPB	7/1/2022 13:14 SMPL
N3356-08	MEY0D0	166	0.0552 -	SFAM01.1	PPB	7/1/2022 13:17 SMPL
N3356-09	MEY0D2	167	0.0554 -	SFAM01.1	PPB	7/1/2022 13:19 SMPL
N3356-10	MEY0C8	-181	-0.0242 -	SFAM01.1	PPB	7/1/2022 13:21 SMPL
N3356-11	MEY0D3	4	0.0181 -	SFAM01.1	PPB	7/1/2022 13:23 SMPL
PB145978BL	PBW597	-76	-0.0002 -	SFAM01.1	PPB	7/1/2022 13:26 SMPL
N3415-01	MESQP9	114	0.0433 -	SFAM01.1	PPB	7/1/2022 13:28 SMPL
N3415-02	MESQQ0	236	0.0712 -	SFAM01.1	PPB	7/1/2022 13:30 SMPL
N3415-04	MESQQ2	207	0.0646 -	SFAM01.1	PPB	7/1/2022 13:32 SMPL
N3415-06	MESQQ4	122	0.0451 -	SFAM01.1	PPB	7/1/2022 13:35 SMPL
N3415-11	MESQQ8	88	0.0373 -	SFAM01.1	PPB	7/1/2022 13:37 SMPL
N3415-12	MESQQ8D	84	0.0364 -	SFAM01.1	PPB	7/1/2022 13:39 SMPL
N3415-13	MESQQ8S	4972	1.1555 -	SFAM01.1	PPB	7/1/2022 13:44 SMPL
CCV076	CCV076	21830	5.0152 -	SFAM01.1	PPB	7/1/2022 13:46 SMPL
CCB076	CCB076	-141	-0.0151 -	SFAM01.1	PPB	7/1/2022 13:49 SMPL