

LB120918


SFAM01.1

INSTRUMENT ID: CV9

Linear

μ Abs.:

4770E



A= 0.0000e+000

B= 2.1272e-004 slope

C= -1.4634e-001 $f = slope \times conc$

Rho= 0.9999684

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	-0.033	-0.033	534	0.000	534					
0.05	0.050					185	0				
0.20	0.200	0.229	0.029	1766	0.0 %	1766					
2.50	2.500	2.531	0.031	12586	0.0 %	12586					
5.00	5.000	4.955	-0.045	23981	0.0 %	23981					
7.50	7.500	7.515	0.015	36018	0.0 %	36018					
10.0	10.000	10.002	0.002	47708	0.0 %	47708					

9/0 D
= 15
- 1
0
0

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	534	-		0 SFAM01.1	PPB	7/1/2022 14:15	Std
	0.2 S01	1766	-		0.2 SFAM01.1	PPB	7/1/2022 14:18	Std
	2.5 S02	12586	-		2.5 SFAM01.1	PPB	7/1/2022 14:20	Std
	5 S03	23981	-		5 SFAM01.1	PPB	7/1/2022 14:22	Std
	7.5 S04	36018	-		7.5 SFAM01.1	PPB	7/1/2022 14:24	Std
	10 S05	47708	-		10 SFAM01.1	PPB	7/1/2022 14:27	Std
ICV022	ICV022	18955	3.8858 -		SFAM01.1	PPB	7/1/2022 14:30	SMPL
ICB022	ICB022	522	-0.0353 -		SFAM01.1	PPB	7/1/2022 14:32	SMPL
CCV077	CCV077	23941	4.9464 -		SFAM01.1	PPB	7/1/2022 14:34	SMPL
CCB077	CCB077	488	-0.0425 -		SFAM01.1	PPB	7/1/2022 14:36	SMPL
PB145998BL	PBS998	623	-0.0138 -		SFAM01.1	PPB	7/1/2022 14:39	SMPL
N3403-01	MC0AA0	753	0.0138 -		SFAM01.1	PPB	7/1/2022 14:41	SMPL
N3403-02	MC0AC7	1069	0.0811 -		SFAM01.1	PPB	7/1/2022 14:43	SMPL
N3403-03	MC0AC7D	1094	0.0864 -		SFAM01.1	PPB	7/1/2022 14:45	SMPL
N3403-04	MC0AC7S	13803	2.7898 -		SFAM01.1	PPB	7/1/2022 14:50	SMPL
N3403-05	MC0AC8	624	-0.0136 -		SFAM01.1	PPB	7/1/2022 14:52	SMPL
N3403-08	MC0AF7	1114	0.0906 -		SFAM01.1	PPB	7/1/2022 14:55	SMPL
CCV078	CCV078	24118	4.9841 -		SFAM01.1	PPB	7/1/2022 14:57	SMPL
CCB078	CCB078	481	-0.044 -		SFAM01.1	PPB	7/1/2022 14:59	SMPL