


LB120922

7470A

INSTRUMENT ID: CV1

μ Abs.:

40213



Linear

A= 0.0000e+000

B= 2.4192e-004

C= -1.2215e-001

Rho= 0.9975922

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	% 1)
0.0	0.000	-0.037	-0.037	350	0.000	350					-
0.2	0.200	0.203	0.003	1342	0.0 %	1342					2
2.5	2.500	2.418	-0.082	10498	0.0 %	10498					-3
5.0	5.000	5.033	0.033	21311	0.0 %	21311					1
7.5	7.500	7.978	0.478	33483	0.0 %	33483					6
10.0	10.000	9.606	-0.394	40213	0.0 %	40213					-4

LB120922 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 01 Jul 2022 16:00:25

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	350	-	0.0000	7470A	PPB	01 Jul 2022 16:04:37	S	Std
0.2 - 1	50.2	1342	-	0.2000	7470A	PPB	01 Jul 2022 16:06:53	S	Std
2.5 - 1	52.5	10498	-	2.5000	7470A	PPB	01 Jul 2022 16:09:10	S	Std
5.0 - 1	55	21311	-	5.0000	7470A	PPB	01 Jul 2022 16:11:27	S	Std
7.5 - 1	57.5	33483	-	7.5000	7470A	PPB	01 Jul 2022 16:13:43	S	Std
10.0 - 1	510	40213	-	10.0000	7470A	PPB	01 Jul 2022 16:19:29	S	Std
ICV23 - 1	ICV23	17161	4.0294	-	7470A	PPB	01 Jul 2022 16:24:49	U	SMPL
ICB23 - 1	ICB23	299	-0.0498	-	7470A	PPB	01 Jul 2022 16:27:04	U	SMPL
CCV79 - 1	CCV79	21020	4.9630	-	7470A	PPB	01 Jul 2022 16:29:19	U	SMPL
CCB79 - 1	CCB79	301	-0.0493	-	7470A	PPB	01 Jul 2022 16:31:34	U	SMPL
CRA - 1	CRA	1322	0.1977	-	7470A	PPB	01 Jul 2022 16:33:49	U	SMPL
HighStd - 1	HighStd	40686	9.7205	-	7470A	PPB	01 Jul 2022 16:36:04	U	SMPL
ChkStd - 1	ChkStd	28885	6.8657	-	7470A	PPB	01 Jul 2022 16:38:19	U	SMPL
PB146005BL - 1	PBW	155	-0.0846	-	7470A	PPB	01 Jul 2022 16:40:35	U	SMPL
PB146005BS - 1	LCSW	17063	4.0057	-	7470A	PPB	01 Jul 2022 16:46:45	U	SMPL
N3535-01 - 1	RW8-20220628	232	-0.0660	-	7470A	PPB	01 Jul 2022 16:49:01	U	SMPL
N3535-01DUP - 1	RW8-20220628DUP	300	-0.0496	-	7470A	PPB	01 Jul 2022 16:51:18	U	SMPL
N3535-01MS - 1	RW8-20220628MS	17485	4.1078	-	7470A	PPB	01 Jul 2022 16:55:54	U	SMPL
N3535-01MSD - 1	RW8-20220628MSD	17463	4.1025	-	7470A	PPB	01 Jul 2022 16:58:10	U	SMPL
N3535-03 - 1	RW8-F-20220628	301	-0.0493	-	7470A	PPB	01 Jul 2022 17:00:25	U	SMPL
CCV80 - 1	CCV80	21050	4.9702	-	7470A	PPB	01 Jul 2022 17:02:41	U	SMPL
CCB80 - 1	CCB80	287	-0.0527	-	7470A	PPB	01 Jul 2022 17:04:56	U	SMPL
N3543-01 - 1	KENS-062822-SB18-GW	360	-0.0351	-	7470A	PPB	01 Jul 2022 17:07:11	U	SMPL
N3543-02 - 1	KENS-062822-SB18-GW	337	-0.0406	-	7470A	PPB	01 Jul 2022 17:09:26	U	SMPL
N3562-19 - 1	FB	334	-0.0413	-	7470A	PPB	01 Jul 2022 17:11:41	U	SMPL
N3563-01 - 1	TRENCH	9512	2.1790	-	7470A	PPB	01 Jul 2022 17:13:57	U	SMPL
N3563-02 - 1	SPOT-3	473	-0.0077	-	7470A	PPB	01 Jul 2022 17:16:12	U	SMPL
N3535-01LX5 - 1		299	-0.0498	-	7470A	PPB	01 Jul 2022 17:18:28	U	SMPL
N3535-01A - 1		17161	4.0294	-	7470A	PPB	01 Jul 2022 17:20:45	U	SMPL
CCV81 - 1	CCV81	19971	4.7092	-	7470A	PPB	01 Jul 2022 17:26:28	U	SMPL
CCB81 - 1	CCB81	283	-0.0537	-	7470A	PPB	01 Jul 2022 17:28:45	U	SMPL