

LB120928


7471B

INSTRUMENT ID: CV2

Linear

 μ Abs.:

53982



A= 0.0000e+000

B= 1.8457e-004

C= 1.2695e-002

Rho= 0.9999820

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	g/d
0.0	0.000	0.005	0.005	-41	0.000	-41					
0.2	0.200	0.195	-0.005	990	0.0 %	990					
2.5	2.500	2.496	-0.004	13457	0.0 %	13457					
5.0	5.000	4.982	-0.018	26926	0.0 %	26926					
7.5	7.500	7.545	0.045	40809	0.0 %	40809					
10.0	10.000	9.976	-0.024	53982	0.0 %	53982					

0-0031

LB120928 INSTRUMENT ID : CV1

Method: 7471B		Operator: Admin		Date of Analysis: 05 Jul 2022 13:59:33					
Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-41	-	0.0000	7471B	PPB	05 Jul 2022 14:05:02	S	Std
0.2 - 1	50.2	990	-	0.2000	7471B	PPB	05 Jul 2022 14:07:18	S	Std
2.5 - 1	52.5	13457	-	2.5000	7471B	PPB	05 Jul 2022 14:09:35	S	Std
5.0 - 1	55	26926	-	5.0000	7471B	PPB	05 Jul 2022 14:11:52	S	Std
7.5 - 1	57.5	40809	-	7.5000	7471B	PPB	05 Jul 2022 14:14:08	S	Std
10.0 - 1	510	53982	-	10.0000	7471B	PPB	05 Jul 2022 14:16:24	S	Std
ICV25 - 1	ICV25	21765	4.0298	-	7471B	PPB	05 Jul 2022 14:19:21	U	SMPL
ICB25 - 1	ICB25	73	0.0262	-	7471B	PPB	05 Jul 2022 14:21:35	U	SMPL
CCV85 - 1	CCV85	26918	4.9809	-	7471B	PPB	05 Jul 2022 14:23:51	U	SMPL
CCB85 - 1	CCB85	56	0.0230	-	7471B	PPB	05 Jul 2022 14:26:06	U	SMPL
CRA - 1	CRA	1019	0.2008	-	7471B	PPB	05 Jul 2022 14:28:22	U	SMPL
HighStd - 1	HighStd	55403	10.2383	-	7471B	PPB	05 Jul 2022 14:30:37	U	SMPL
ChkStd - 1	ChkStd	39389	7.2826	-	7471B	PPB	05 Jul 2022 14:32:52	U	SMPL
PB146015BL - 1	PBS	-165	-0.0178	-	7471B	PPB	05 Jul 2022 14:35:10	U	SMPL
PB146015BS - 1	LCSS	19886	3.6830	-	7471B	PPB	05 Jul 2022 14:37:27	U	SMPL
N3557-01 - 1	OR-02-063022	1556	0.2999	-	7471B	PPB	05 Jul 2022 14:39:43	U	SMPL
N3557-01DUP - 1	OR-02-063022DUP	1538	0.2966	-	7471B	PPB	05 Jul 2022 14:49:28	U	SMPL
N3557-01MS - 1	OR-02-063022MS	21845	4.0446	-	7471B	PPB	05 Jul 2022 14:51:44	U	SMPL
N3557-01MSD - 1	OR-02-063022MSD	21625	4.0040	-	7471B	PPB	05 Jul 2022 14:54:01	U	SMPL
N3558-01 - 1	72-11998	2656	0.5029	-	7471B	PPB	05 Jul 2022 14:56:17	U	SMPL
CCV86 - 1	CCV86	26774	4.9543	-	7471B	PPB	05 Jul 2022 14:58:33	U	SMPL
CCB86 - 1	CCB86	68	0.0252	-	7471B	PPB	05 Jul 2022 15:00:48	U	SMPL
N3564-01 - 1	WC-14B-11-1	13402	2.4863	-	7471B	PPB	05 Jul 2022 15:03:05	U	SMPL
N3564-06 - 1	WC-14B-11-2	9985	1.8556	-	7471B	PPB	05 Jul 2022 15:05:20	U	SMPL
N3564-11 - 1	WC-14B-11-3	13890	2.5763	-	7471B	PPB	05 Jul 2022 15:07:35	U	SMPL
N3564-16 - 1	WC-16-2-6	712407	HIGH	-	7471B	PPB	05 Jul 2022 15:09:51	X	
N3564-21 - 1	WC-16-2-5	265510	49.0171	-	7471B	PPB	05 Jul 2022 15:11:41	U	SMPL
N3564-26 - 1	WC-16-2-4	326857	60.3398	-	7471B	PPB	05 Jul 2022 15:14:07	U	SMPL
N3564-31 - 1	WC-16-2-3	32378	5.9886	-	7471B	PPB	05 Jul 2022 15:16:31	U	SMPL
N3567-01 - 1	OK-01-07012022	2300	0.4372	-	7471B	PPB	05 Jul 2022 15:18:59	U	SMPL
N3568-01 - 1	TR-05-07012022	771	0.1550	-	7471B	PPB	05 Jul 2022 15:21:18	U	SMPL
N3571-01 - 1	VB16128	3355	0.6319	-	7471B	PPB	05 Jul 2022 15:23:34	U	SMPL
CCV87 - 1	CCV87	27082	5.0111	-	7471B	PPB	05 Jul 2022 15:25:49	U	SMPL
CCB87 - 1	CCB87	-12	0.0105	-	7471B	PPB	05 Jul 2022 15:28:04	U	SMPL
N3572-01 - 1	RBR-200033	3025	0.5710	-	7471B	PPB	05 Jul 2022 15:30:21	U	SMPL
PB146016BL - 1	PBS	-26	0.0079	-	7471B	PPB	05 Jul 2022 15:32:36	U	SMPL
PB146016BS - 1	LCSS	19719	3.6522	-	7471B	PPB	05 Jul 2022 15:34:51	U	SMPL
N3562-01 - 1	SB-7A	2172	0.4136	-	7471B	PPB	05 Jul 2022 15:37:07	U	SMPL
N3562-01DUP - 1	SB-7ADUP	1950	0.3726	-	7471B	PPB	05 Jul 2022 15:39:24	U	SMPL
N3562-01MS - 1	SB-7AMS	21677	4.0136	-	7471B	PPB	05 Jul 2022 15:41:40	U	SMPL
N3562-01MSD - 1	SB-7AMSD	21918	4.0580	-	7471B	PPB	05 Jul 2022 15:43:56	U	SMPL
N3562-02 - 1	SB-7B	15058	2.7919	-	7471B	PPB	05 Jul 2022 15:46:13	U	SMPL
N3562-03 - 1	SB-8A	2356	0.4475	-	7471B	PPB	05 Jul 2022 15:48:31	U	SMPL
N3562-04 - 1	SB-8B	3300	0.6218	-	7471B	PPB	05 Jul 2022 15:50:47	U	SMPL
CCV88 - 1	CCV88	26868	4.9716	-	7471B	PPB	05 Jul 2022 15:53:02	U	SMPL
CCB88 - 1	CCB88	-100	-0.0058	-	7471B	PPB	05 Jul 2022 15:55:17	U	SMPL
N3562-05 - 1	SB-9A	3474	0.6539	-	7471B	PPB	05 Jul 2022 15:57:33	U	SMPL
N3562-06 - 1	SB-9B	794	0.1592	-	7471B	PPB	05 Jul 2022 15:59:49	U	SMPL
N3562-07 - 1	SB-10A	12833	2.3812	-	7471B	PPB	05 Jul 2022 16:02:04	U	SMPL
N3562-08 - 1	SB-10B	29598	5.4755	-	7471B	PPB	05 Jul 2022 16:04:20	U	SMPL
N3562-09 - 1	SB-11A	4812	0.9008	-	7471B	PPB	05 Jul 2022 16:06:36	U	SMPL
N3562-10 - 1	SB-11B	279	0.0642	-	7471B	PPB	05 Jul 2022 16:08:53	U	SMPL
N3562-11 - 1	SB-12A	2408	0.4571	-	7471B	PPB	05 Jul 2022 16:11:09	U	SMPL
N3562-12 - 1	SB-12B	5542	1.0356	-	7471B	PPB	05 Jul 2022 16:13:26	U	SMPL
N3562-13 - 1	D62922	2839	0.5367	-	7471B	PPB	05 Jul 2022 16:15:42	U	SMPL
CCV89 - 1	CCV89	27929	5.1675	-	7471B	PPB	05 Jul 2022 16:17:58	U	SMPL
CCB89 - 1	CCB89	-84	-0.0028	-	7471B	PPB	05 Jul 2022 16:20:13	U	SMPL
N3557-01LX5 - 1		248	0.0585	-	7471B	PPB	05 Jul 2022 16:22:30	U	SMPL
N3557-01A - 1		22868	4.2334	-	7471B	PPB	05 Jul 2022 16:24:46	U	SMPL
N3562-01LX5 - 1		309	0.0697	-	7471B	PPB	05 Jul 2022 16:27:01	U	SMPL
N3562-01A - 1		22028	4.0783	-	7471B	PPB	05 Jul 2022 16:29:17	U	SMPL
N3564-16DLX10 - 1		139061	25.6788	-	7471B	PPB	05 Jul 2022 16:31:32	U	SMPL
N3564-21DLX10 - 1		30361	5.6163	-	7471B	PPB	05 Jul 2022 16:33:49	U	SMPL
N3564-26DLX10 - 1		37640	6.9598	-	7471B	PPB	05 Jul 2022 16:36:08	U	SMPL
N3564-16DL2X50 - 1		29560	5.4685	-	7471B	PPB	05 Jul 2022 16:38:26	U	SMPL
CCV90 - 1	CCV90	27218	5.0362	-	7471B	PPB	05 Jul 2022 16:40:44	U	SMPL
CCB90 - 1	CCB90	60	0.0238	-	7471B	PPB	05 Jul 2022 16:43:02	U	SMPL