



Analytical Summary Report

Analysis Method: 9040C
Parameter: Corrosivity
Run Number: LB120937

Analyst By : JIGNESH
Supervisor Review By : mohammad
Slope : 99.2
pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2905
BUFFER PH 7.00 GREEN 1PINT PK6	W2906
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.9	4.01	07/06/2022	09:40
2	CAL2	1	Water	NA	NA	20.9	7.00	07/06/2022	09:41
3	CAL3	1	Water	NA	NA	20.9	10.03	07/06/2022	09:42
4	ICV	1	Water	NA	NA	20.8	6.99	07/06/2022	09:45
5	CCV1	1	Water	NA	NA	20.8	2.02	07/06/2022	10:15
6	N3575-01	1	Water	NA	NA	23.6	7.57	07/06/2022	10:20
7	N3575-01DUP	1	Water	NA	NA	23.7	7.58	07/06/2022	10:21
8	N3583-02	1	Water	NA	NA	23.2	8.44	07/06/2022	10:25
9	N3583-03	1	Water	NA	NA	21.7	7.20	07/06/2022	10:26
10	CCV2	1	Water	NA	NA	20.9	12.03	07/06/2022	10:30

WORKLIST(Hardcopy Internal Chain)

NR 120037

WorkList Name : corrosivity n3580 w WorkList ID : 161125 Department : Wet-Chemistry Date : 07-06-2022 09:33:53

Due Date	Matrix	Sample	Test	Raw Sample			Collect Date	Method
				Preservative	Customer	Storage Location		
07/01/2022	Water	N3575-01	Corrosivity	Cool 4 deg C	DVIR01	F11 WC-4	07/01/2022	9040C
07/05/2022	Water	N3583-02	Corrosivity	Cool 4 deg C	PSEG03	G11 307	07/05/2022	9040C
07/05/2022	Water	N3583-03	Corrosivity	Cool 4 deg C	PSEG03	G11 311	07/05/2022	9040C

Date/Time 07/06/22 09:35
 Raw Sample Received by: P (WC)
 Raw Sample Relinquished by: [Signature]

Date/Time 07/06/22 12:00
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: P (WC)