



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : mohammad

Run Number: LB120938

Slope : 99.2

BalanceID: WC SC-6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2905
BUFFER PH 7.00 GREEN 1PINT PK6	W2906
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.9	4.01	07/06/2022	09:40
2	CAL2	1	Water	NA	NA	20.9	7.00	07/06/2022	09:41
3	CAL3	1	Water	NA	NA	20.9	10.03	07/06/2022	09:42
4	ICV	1	Water	NA	NA	20.8	6.99	07/06/2022	09:45
5	CCV1	1	Water	NA	NA	20.7	2.01	07/06/2022	10:33
6	N3581-01	1	Solid	20.03	20	25.7	6.74	07/06/2022	10:35
7	N3581-01DUP	1	Solid	20.04	20	25.8	6.75	07/06/2022	10:36
8	CCV2	1	Water	NA	NA	20.6	12.03	07/06/2022	10:40

WORKLIST(Hardcopy Internal Chain)

120938

WorkList Name : ph-070622

WorkList ID : 161114

Department : Wet-Chemistry

Date : 07-06-2022 08:04:00

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
07/05/2022	Solid	N3581-01	pH	Cool 4 deg C	PSEG05	N11	AU-6-070522	07/05/2022	9045D

09/35

Date/Time 07/06/22 12:00
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]

07/06/22 12:00

Date/Time 07/06/22 12:00
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]