


68120939

7470A

INSTRUMENT ID: CV1

μ Abs.:

42843



Linear

A= 0.0000e+000

B= 2.3573e-004

C= 4.1376e-002

Rho= 0.9995184

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	V/O
0.0	0.000	0.037	0.037	-20	0.000	-20					
0.2	0.200	0.219	0.019	753	0.0 %	753					9
2.5	2.500	2.474	-0.026	10318	0.0 %	10318					-1
5.0	5.000	5.060	0.060	21289	0.0 %	21289					1
7.5	7.500	7.270	-0.230	30665	0.0 %	30665					-3
10.0	10.000	10.141	0.141	42843	0.0 %	42843					1

LB120939 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 06 Jul 2022 11:30:19

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-20	-	0.0000	7470A	PPB	06 Jul 2022 11:31:25	S	Std
0.2 - 1	50.2	753	-	0.2000	7470A	PPB	06 Jul 2022 11:33:42	S	Std
2.5 - 1	52.5	10318	-	2.5000	7470A	PPB	06 Jul 2022 11:35:59	S	Std
5.0 - 1	55	21289	-	5.0000	7470A	PPB	06 Jul 2022 11:38:16	S	Std
7.5 - 1	57.5	30665	-	7.5000	7470A	PPB	06 Jul 2022 11:43:06	S	Std
10.0 - 1	5N	42843	-	10.0000	7470A	PPB	06 Jul 2022 11:45:21	S	Std
ICV26 - 1	ICV26	16517	3.9350	-	7470A	PPB	06 Jul 2022 11:48:13	U	SMPL
ICB26 - 1	ICB26	-66	0.0258	-	7470A	PPB	06 Jul 2022 11:50:28	U	SMPL
CCV91 - 1	CCV91	20195	4.8020	-	7470A	PPB	06 Jul 2022 11:52:43	U	SMPL
CCB91 - 1	CCB91	-54	0.0286	-	7470A	PPB	06 Jul 2022 11:54:58	U	SMPL
CRA - 1	CRA	698	0.2059	-	7470A	PPB	06 Jul 2022 11:57:13	U	SMPL
HighStd - 1	HighStd	43177	10.2196	-	7470A	PPB	06 Jul 2022 11:59:28	U	SMPL
ChkStd - 1	ChkStd	28261	6.7034	-	7470A	PPB	06 Jul 2022 12:01:44	U	SMPL
PB146032BL - 1	PBW	-102	0.0173	-	7470A	PPB	06 Jul 2022 12:04:01	U	SMPL
PB146032BS - 1	LCSW	14405	3.4371	-	7470A	PPB	06 Jul 2022 12:06:18	U	SMPL
N3564-02 - 1	WC-14B-11-1	77	0.0595	-	7470A	PPB	06 Jul 2022 12:08:34	U	SMPL
N3564-07 - 1	WC-14B-11-2	37	0.0501	-	7470A	PPB	06 Jul 2022 12:10:50	U	SMPL
N3564-12 - 1	WC-14B-11-3	21	0.0463	-	7470A	PPB	06 Jul 2022 12:13:06	U	SMPL
N3564-17 - 1	WC-16-2-6	223	0.0939	-	7470A	PPB	06 Jul 2022 12:15:23	U	SMPL
N3564-22 - 1	WC-16-2-5	393	0.1340	-	7470A	PPB	06 Jul 2022 12:17:38	U	SMPL
CCV92 - 1	CCV92	21053	5.0042	-	7470A	PPB	06 Jul 2022 12:19:53	U	SMPL
CCB92 - 1	CCB92	-67	0.0256	-	7470A	PPB	06 Jul 2022 12:22:09	U	SMPL
N3564-27 - 1	WC-16-2-4	463	0.1505	-	7470A	PPB	06 Jul 2022 12:24:24	U	SMPL
N3564-32 - 1	WC-16-2-3	154	0.0777	-	7470A	PPB	06 Jul 2022 12:26:40	U	SMPL
N3566-01 - 1	LGS-PAINT-0701	-8	0.0395	-	7470A	PPB	06 Jul 2022 12:28:55	U	SMPL
N3571-02 - 1	VB16128	-16	0.0376	-	7470A	PPB	06 Jul 2022 12:31:11	U	SMPL
N3571-02DUP - 1	VB16128DUP	-24	0.0357	-	7470A	PPB	06 Jul 2022 12:33:26	U	SMPL
N3571-02MS - 1	VB16128MS	14414	3.4392	-	7470A	PPB	06 Jul 2022 12:35:43	U	SMPL
N3571-02MSD - 1	VB16128MSD	14688	3.5038	-	7470A	PPB	06 Jul 2022 12:37:59	U	SMPL
N3571-02LX5 - 1		-104	0.0169	-	7470A	PPB	06 Jul 2022 12:40:16	U	SMPL
N3571-02A - 1		16377	3.9020	-	7470A	PPB	06 Jul 2022 12:42:33	U	SMPL
PB145990TB - 1	PB145990TB	-66	0.0258	-	7470A	PPB	06 Jul 2022 12:44:49	U	SMPL
CCV93 - 1	CCV93	21263	5.0537	-	7470A	PPB	06 Jul 2022 12:47:05	U	SMPL
CCB93 - 1	CCB93	-62	0.0268	-	7470A	PPB	06 Jul 2022 12:49:20	U	SMPL