



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Guerda

Parameter: Flash Point

Supervisor Review By: mohammad

Run Number: LB120943

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2743

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.3	07/06/2022	09:25
2	N3575-01		1	15	100.00	>212.0	>212.0	07/06/2022	09:50
3	N3575-01DUP		1	15	100.00	>212.0	>212.0	07/06/2022	10:10
4	N3583-02		1	18	100.00	>212.0	>212.0	07/06/2022	10:35
5	N3583-03		1	17	100.00	>212.0	>212.0	07/06/2022	10:58
6	N3586-06		1	16	100.00	>212.0	>212.0	07/06/2022	11:20
7	N3586-08		1	18	100.00	>212.0	>212.0	07/06/2022	11:45

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB 120943

WorkList Name : FLASHPOINT-07-06-22

WorkList ID : 161120

Department : Wet-Chemistry

Date : 07-06-2022 09:00:57

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
07/07/2022	Water	N3575-01	Flash Point	Cool 4 deg C	DVIR01	F11	WC-4	07/01/2022	1010B
07/12/2022	Water	N3583-02	Flash Point	Cool 4 deg C	PSEG03	G11	307	07/05/2022	1010B
07/12/2022	Water	N3583-03	Flash Point	Cool 4 deg C	PSEG03	G11	311	07/05/2022	1010B
07/08/2022	Water	N3586-06	Flash Point	Cool 4 deg C	PSEG03	F11	41043	07/05/2022	1010B
07/08/2022	Water	N3586-08	Flash Point	Cool 4 deg C	PSEG03	F11	31187	07/05/2022	1010B

Date/Time 07/06/22 09:10
 Raw Sample Received by: SM (WC)
 Raw Sample Relinquished by: KS (WC)

Date/Time 07/06/22 11:50
 Raw Sample Received by: KS (WC)
 Raw Sample Relinquished by: SM (WC)