



Analytical Summary Report

Analysis Method: 365.3

ANALYST: Iwona

Parameter: Phosphorus, Total

SUPERVISOR REVIEW BY: JIGNESH

Run Number: LB120944

BALANCE ID: WC SC-4

Reagent/Standard	Lot/Log #
calibration std. phosphate 1 ppm	WP97743
calibration std. phosphate 0.5 ppm	WP97742
calibration std. phosphate 0.3 ppm	WP97741
calibration std. phosphate 0.1 ppm	WP97740
calibration std. phosphate 0.05 ppm	WP97739
calibration std. 0 ppm	WP97738
phosphate CCV std.	WP97744
Combined reagent	WP97770
Phenolphthalein indicator	WP96861
Sodium hydroxide, 1N	WP95405
Phosphate ICV-LCS Std	WP97745
Phosphate LOD-MDL Std 0.025ppm	WP97759

Intercept: -0.0015

Slope: 0.6497

Regression: 0.999945

Seq	Lab ID	True Value (mg/L)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	Result (mg/L)	%D	AnalDate	AnalTime
1	CAL1	0.00	1	50	50	0.000	0.002		07/06/2022	13:25
2	CAL2	0.05	1	50	50	0.031	0.05	0	07/06/2022	13:25
3	CAL3	0.10	1	50	50	0.066	0.104	-4	07/06/2022	13:26
4	CAL4	0.30	1	50	50	0.189	0.293	2.3	07/06/2022	13:26
5	CAL5	0.50	1	50	50	0.322	0.498	0.4	07/06/2022	13:27
6	CAL6	1.00	1	50	50	0.650	1.003	-0.3	07/06/2022	13:27



Analytical Summary Report

Analysis Method: 365.3

ANALYST: Iwona

Parameter: Phosphorus, Total

SUPERVISOR REVIEW BY: JIGNESH

Run Number: LB120944

BALANCE ID: WC SC-4

Seq	Lab ID	True Value (mg/l)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	Result (mg/L)	AnalDate	AnalTime
1	ICV	0.50	1	50	50	0.322	0.498	07/06/2022	13:28
2	ICB		1	50	50	0.001	0.004	07/06/2022	13:28
3	CCV1	0.50	1	50	50	0.328	0.507	07/06/2022	13:29
4	CCB1		1	50	50	0.003	0.007	07/06/2022	13:29
5	PB146047BL		1	1	50	0.002	0.005	07/06/2022	13:30
6	PB146047BS	0.50	1	1	50	0.326	0.504	07/06/2022	13:30
7	N3214-03		1	1	50	0.017	0.028	07/06/2022	13:31
8	CCV2	0.50	1	50	50	0.325	0.503	07/06/2022	13:31
9	CCB2		1	50	50	0.001	0.004	07/06/2022	13:32