



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB120966

Analyst By : Guerda

Supervisor Review By : jignesh

Slope : 98.1

pH Meter ID : WC pH meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2905
BUFFER PH 7.00 GREEN 1PINT PK6	W2906
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	24.3	4.01	07/07/2022	09:00
2	CAL2	1	Water	NA	NA	21.8	7.00	07/07/2022	09:08
3	CAL3	1	Water	NA	NA	20.7	10.05	07/07/2022	09:16
4	ICV	1	Water	NA	NA	20.3	6.99	07/07/2022	09:18
5	CCV1	1	Water	NA	NA	21.8	2.00	07/07/2022	09:22
6	N3621-02	1	Water	NA	NA	20.3	6.89	07/07/2022	09:28
7	N3621-02DUP	1	Water	NA	NA	20.4	6.90	07/07/2022	09:30
8	CCV2	1	Water	NA	NA	21.8	12.03	07/07/2022	09:36

WORKLIST(Hardcopy Internal Chain)

VB 120988

WorkList Name : ph-070722 WorkList ID : 161151 Department : Wet-Chemistry Date : 07-07-2022 07:41:55

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
07/06/2022	Water	N3621-02	pH		Cool 4 deg C	TETR06	G11	TT-047-IDWGW-20220706	07/06/2022 9040C

Date/Time 7/7/22 8:30
 Raw Sample Received by: SP (nc)
 Raw Sample Relinquished by: SP (nc)

Date/Time 7/7/22 10:00
 Raw Sample Received by: SP (nc)
 Raw Sample Relinquished by: SP (nc)