


LB120968

SFAM01.1

INSTRUMENT ID: CV1

μ Abs.:

4590E



Linear

A= 0.0000e+000

B= 2.1922e-004 slope

C= 1.3529e-002 y-intercept

Rho= 0.9999016

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/20
0.00	0.000	0.019	0.019	23	0.000	23					=
0.05	0.050					185	0				78 m8
0.20	0.200	0.215	0.015	918	0.0 %	918					0 21112
2.50	2.500	2.499	-0.001	11340	0.0 %	11340					0
5.00	5.000	4.986	-0.014	22683	0.0 %	22683					-1
7.50	7.500	7.403	-0.097	33710	0.0 %	33710					1
10.0	10.000	10.078	0.078	45908	0.0 %	45908					

LB120968
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	23	-		0 SFAM01.1	PPB	7/7/2022 14:03	Std
	0.2 S01	918	-		0.2 SFAM01.1	PPB	7/7/2022 14:05	Std
	2.5 S02	11340	-		2.5 SFAM01.1	PPB	7/7/2022 14:07	Std
	5 S03	22683	-		5 SFAM01.1	PPB	7/7/2022 14:12	Std
	7.5 S04	33710	-		7.5 SFAM01.1	PPB	7/7/2022 14:15	Std
	10 S05	45908	-		10 SFAM01.1	PPB	7/7/2022 14:17	Std
ICV028	ICV028	18296	4.0244 -		SFAM01.1	PPB	7/7/2022 14:20	SMPL
ICB028	ICB028	-8	0.0118 -		SFAM01.1	PPB	7/7/2022 14:22	SMPL
CCV097	CCV097	22369	4.9173 -		SFAM01.1	PPB	7/7/2022 14:24	SMPL
CCB097	CCB097	-6	0.0122 -		SFAM01.1	PPB	7/7/2022 14:26	SMPL
PB146087BL	PBW087	103	0.0361 -		SFAM01.1	PPB	7/7/2022 14:29	SMPL
N3519-01	MH0AC6	116	0.039 -		SFAM01.1	PPB	7/7/2022 14:31	SMPL
N3519-02	MH0AC8	139	0.044 -		SFAM01.1	PPB	7/7/2022 14:33	SMPL
N3519-03	MH0AD0	1792	0.4064 -		SFAM01.1	PPB	7/7/2022 14:35	SMPL
N3519-04	MH0AD0D	1609	0.3663 -		SFAM01.1	PPB	7/7/2022 14:38	SMPL
N3519-05	MH0AD0S	5910	1.3091 -		SFAM01.1	PPB	7/7/2022 14:40	SMPL
N3519-06	MH0AD2	62	0.0271 -		SFAM01.1	PPB	7/7/2022 14:42	SMPL
N3519-07	MH0AD4	126	0.0412 -		SFAM01.1	PPB	7/7/2022 14:45	SMPL
N3519-08	MH0AD6	128	0.0416 -		SFAM01.1	PPB	7/7/2022 14:47	SMPL
N3519-09	MH0AD8	110	0.0376 -		SFAM01.1	PPB	7/7/2022 14:49	SMPL
PB146088BL	PBW088	119	0.0396 -		SFAM01.1	PPB	7/7/2022 14:51	SMPL
N3520-01	MH0AC7	110	0.0376 -		SFAM01.1	PPB	7/7/2022 14:54	SMPL
N3520-02	MH0AC9	110	0.0376 -		SFAM01.1	PPB	7/7/2022 14:56	SMPL
N3520-03	MH0AD1	93	0.0339 -		SFAM01.1	PPB	7/7/2022 14:58	SMPL
N3520-04	MH0AD1D	79	0.0308 -		SFAM01.1	PPB	7/7/2022 15:00	SMPL
N3520-05	MH0AD1S	3869	0.8617 -		SFAM01.1	PPB	7/7/2022 15:03	SMPL
N3520-06	MH0AD3	148	0.046 -		SFAM01.1	PPB	7/7/2022 15:05	SMPL
N3520-07	MH0AD5	119	0.0396 -		SFAM01.1	PPB	7/7/2022 15:07	SMPL
N3520-08	MH0AD7	128	0.0416 -		SFAM01.1	PPB	7/7/2022 15:09	SMPL
N3520-09	MH0AD9	89	0.033 -		SFAM01.1	PPB	7/7/2022 15:12	SMPL
CCV098	CCV098	22616	4.9714 -		SFAM01.1	PPB	7/7/2022 15:14	SMPL
CCB098	CCB098	190	0.0552 -		SFAM01.1	PPB	7/7/2022 15:16	SMPL