



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: Guerda

Parameter: Turbidity

SUPERVISOR REVIEW BY: mohammad

Run Number: LB120988

Reagent/Standard	Lot/Log #
TURBIDITY LOD STD, 0.5NTU	WP97839
Turbidity Calibration std, 0NTU	WP97838
Turbidity Calibration std, 20NTU	WP97834
Turbidity Calibration std, 40NTU	WP97833
Turbidity Calibration std, 1NTU	WP97837
Turbidity Calibration std, 5NTU	WP97836
Turbidity Calibration - CCV std, 10 NTU	WP97840

Intercept: 0.0385

Slope: 0.9343

Regression: 0.9998

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (mg/l)	%D	Anal Date	Anal Time
1	CAL1	0	1	0.146	0.12		07/08/2022	16:40
2	CAL2	1	1	1.152	1.19	-19.2	07/08/2022	16:43
3	CAL3	5	1	4.460	4.73	5.4	07/08/2022	16:46
4	CAL4	10	1	9.580	10.21	-2.1	07/08/2022	16:49
5	CAL5	20	1	18.300	19.55	2.3	07/08/2022	16:52
6	CAL6	40	1	37.600	40.20	-0.5	07/08/2022	16:55



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: Guerda

Parameter: Turbidity

SUPERVISOR REVIEW BY: mohammad

Run Number: LB120988

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	9.460	10.084	07/08/2022	16:58
2	ICB		1	0.140	0.109	07/08/2022	17:01
3	CCV1	10	1	9.540	10.170	07/08/2022	17:04
4	CCB1		1	0.138	0.106	07/08/2022	17:07
5	LB120988BL		1	0.156	0.126	07/08/2022	17:10
6	N3214-12		1	0.520	0.515	07/08/2022	17:13
7	N3642-01		1	2.580	2.720	07/08/2022	17:16
8	N3642-01DUP		1	2.570	2.710	07/08/2022	17:19
9	N3642-03		1	1.040	1.072	07/08/2022	17:22
10	N3642-05		1	3.430	3.630	07/08/2022	17:25
11	N3642-07		1	2.620	2.763	07/08/2022	17:28
12	CCV2	10	1	9.230	9.838	07/08/2022	17:31
13	CCB2		1	0.144	0.113	07/08/2022	17:34

WORKLIST(Hardcopy Internal Chain)

LB 120488

WorkList Name : TURB-070822 WorkList ID : 161217 Department : Wet-Chemistry Date : 07-08-2022 12:50:27

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
07/09/2022	Water	N3642-01	Turbidity	Cool 4 deg C	LOCK01	G11	MW-1	07/07/2022	SM2130 B
07/09/2022	Water	N3642-03	Turbidity	Cool 4 deg C	LOCK01	G11	MW-2	07/07/2022	SM2130 B
07/09/2022	Water	N3642-05	Turbidity	Cool 4 deg C	LOCK01	G11	MW-3	07/07/2022	SM2130 B
07/09/2022	Water	N3642-07	Turbidity	Cool 4 deg C	LOCK01	G11	MW-4	07/07/2022	SM2130 B

Date/Time 7/8/22 16:35
 Raw Sample Received by: SM (rec)
 Raw Sample Relinquished by: SP1051

Date/Time 7/8/22 17:40
 Raw Sample Received by: SP1051
 Raw Sample Relinquished by: SP1051