



Analytical Summary Report

Analysis Method: 180.1

ANALYST: Guerda

Parameter: Turbidity

SUPERVISOR REVIEW BY: mohammad

Run Number: LB120989

Reagent/Standard	Lot/Log #
TURBIDITY LOD STD, 0.5NTU	WP97839
Turbidity Calibration std, 0NTU	WP97838
Turbidity Calibration std, 20NTU	WP97834
Turbidity Calibration std, 40NTU	WP97833
Turbidity Calibration std, 1NTU	WP97837
Turbidity Calibration std, 5NTU	WP97836
Turbidity Calibration - CCV std, 10 NTU	WP97840

Intercept: 0.0385

Slope: 0.9343

Regression: 0.9998

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (mg/l)	%D	Anal Date	Anal Time
1	CAL1	0	1	0.146	0.12		07/09/2022	16:40
2	CAL2	1	1	1.152	1.19	-19.2	07/08/2022	16:43
3	CAL3	5	1	4.460	4.73	5.4	07/08/2022	16:46
4	CAL4	10	1	9.580	10.21	-2.1	07/08/2022	16:49
5	CAL5	20	1	18.300	19.55	2.3	07/08/2022	16:52
6	CAL6	40	1	37.600	40.20	-0.5	07/08/2022	16:55



Analytical Summary Report

Analysis Method: 180.1

ANALYST: Guerda

Parameter: Turbidity

SUPERVISOR REVIEW BY: mohammad

Run Number: LB120989

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	9.460	10.084	07/08/2022	16:58
2	ICB		1	0.140	0.109	07/08/2022	17:01
3	CCV1	10	1	9.540	10.170	07/08/2022	17:04
4	CCB1		1	0.138	0.106	07/08/2022	17:07
5	LB120989BL		1	0.156	0.126	07/08/2022	17:10
6	N3214-09		1	0.520	0.515	07/08/2022	17:13
7	CCV2	10	1	9.230	9.838	07/08/2022	17:31
8	CCB2		1	0.144	0.113	07/08/2022	17:34