



Analytical Summary Report

Analysis Method: SM 4500-H B

Parameter: pH

Run Number: LB120997

Analyst By : JIGNESH

Supervisor Review By : mohammad

Slope : 99.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2905
BUFFER PH 7.00 GREEN 1PINT PK6	W2906
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	07/08/2022	17:15
2	CAL2	1	Water	NA	NA	20.7	7.01	07/08/2022	17:16
3	CAL3	1	Water	NA	NA	20.7	10.02	07/08/2022	17:17
4	ICV	1	Water	NA	NA	20.8	6.99	07/08/2022	17:20
5	CCV1	1	Water	NA	NA	20.8	2.01	07/08/2022	17:22
6	N3642-01	1	Water	NA	NA	20.1	6.33	07/08/2022	17:25
7	N3642-01DUP	1	Water	NA	NA	20.2	6.34	07/08/2022	17:26
8	N3642-03	1	Water	NA	NA	20.7	6.66	07/08/2022	17:30
9	N3642-05	1	Water	NA	NA	20.3	6.78	07/08/2022	17:31
10	N3642-07	1	Water	NA	NA	20.2	6.83	07/08/2022	17:33
11	CCV2	1	Water	NA	NA	20.2	12.02	07/08/2022	17:35

WORKLIST(Hardcopy Internal Chain)

13120997

WorkList Name : ph n3642 WorkList ID : 161230 Department : Wet-Chemistry Date : 07-08-2022 17:10:07

Due Date	Matrix	Sample	Test	Raw Sample			Collect Date	Method
				Preservative	Customer	Storage Location		
07/07/2022	Water	N3642-01	pH C	Cool 4 deg C	LOCK01	G11 MW-1	07/07/2022	SM 4500-H B
07/07/2022	Water	N3642-03	pH C	Cool 4 deg C	LOCK01	G11 MW-2	07/07/2022	SM 4500-H B
07/07/2022	Water	N3642-05	pH C	Cool 4 deg C	LOCK01	G11 MW-3	07/07/2022	SM 4500-H B
07/07/2022	Water	N3642-07	pH C	Cool 4 deg C	LOCK01	G11 MW-4	07/07/2022	SM 4500-H B

Date/Time 07/08/22 17:11
 Raw Sample Received by: JPLWC
 Raw Sample Relinquished by: JPLWC

Date/Time 07/08/22 18:30
 Raw Sample Received by: JPLWC
 Raw Sample Relinquished by: JPLWC