



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : mohammad

Run Number: LB121000

Slope : 99.2

BalanceID: WC SC-6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2905
BUFFER PH 7.00 GREEN 1PINT PK6	W2906
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	07/08/2022	17:15
2	CAL2	1	Water	NA	NA	20.7	7.01	07/08/2022	17:16
3	CAL3	1	Water	NA	NA	20.7	10.02	07/08/2022	17:17
4	ICV	1	Water	NA	NA	20.8	6.99	07/08/2022	17:20
5	CCV1	1	Water	NA	NA	20.1	2.02	07/08/2022	17:40
6	N3648-02	1	Solid	20.03	20	24.5	7.21	07/08/2022	17:42
7	N3648-02DUP	1	Solid	20.04	20	24.6	7.22	07/08/2022	17:43
8	N3648-05	1	Solid	20.02	20	23.7	8.32	07/08/2022	17:45
9	CCV2	1	Water	NA	NA	20.8	12.03	07/08/2022	17:46

WORKLIST(Hardcopy Internal Chain)

B121000

WorkList Name : corrosivity n3648 WorkList ID : 161233 Department : Wet-Chemistry Date : 07-08-2022 17:25:10

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample		Collect Date	Method
						Storage	Location		
07/08/2022	Solid	N3648-02	Corrosivity	Cool 4 deg C	PSEG03	G11	CMR-TP1-0708	07/08/2022	9045D
07/08/2022	Solid	N3648-05	Corrosivity	Cool 4 deg C	PSEG03	G11	CMR-TP2-0708	07/08/2022	9045D

Date/Time 07/08/22 17:26
 Raw Sample Received by: JPL/ACC
 Raw Sample Relinquished by: JPL/ACC

Date/Time 07/08/22 18:30
 Raw Sample Received by: JPL/ACC
 Raw Sample Relinquished by: JPL/ACC