



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB121005

Slope : 99.2

BalanceID: WC SC-6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2905
BUFFER PH 7.00 GREEN 1PINT PK6	W2906
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.6	4.01	07/01/2022	15:15
2	CAL2	1	Water	NA	NA	20.6	7.00	07/01/2022	15:16
3	CAL3	1	Water	NA	NA	20.7	10.03	07/01/2022	15:20
4	ICV	1	Water	NA	NA	20.7	6.99	07/01/2022	15:22
5	CCV1	1	Water	NA	NA	20.7	2.01	07/01/2022	15:25
6	N3564-02	1	Solid	20.02	20	24.9	7.43	07/01/2022	15:26
7	N3564-02DUP	1	Solid	20.03	20	24.8	7.42	07/01/2022	15:27
8	N3564-07	1	Solid	20.04	20	25.1	7.81	07/01/2022	15:33
9	N3564-12	1	Solid	20.01	20	24.8	7.90	07/01/2022	15:35
10	N3564-17	1	Solid	20.04	20	25.2	10.15	07/01/2022	15:40
11	N3564-22	1	Solid	20.02	20	24.8	8.21	07/01/2022	15:46
12	N3564-27	1	Solid	20.03	20	24.8	8.21	07/01/2022	15:50
13	N3564-32	1	Solid	20.04	20	24.1	10.17	07/01/2022	16:00
14	N3571-02	1	Solid	20.01	20	26.5	8.01	07/01/2022	16:10
15	CCV2	1	Water	NA	NA	20.8	12.02	07/01/2022	16:15