

LB121007


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

55241



A= 0.0000e+000

B= 1.8122e-004

C= 3.5238e-002

Rho= 0.9999409

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	9/00
0.0	0.000	0.032	0.032	-18	0.000	-18					
0.2	0.200	0.226	0.026	1051	0.0 %	1051					13
2.5	2.500	2.475	-0.025	13466	0.0 %	13466					-1
5.0	5.000	4.928	-0.072	26998	0.0 %	26998					-1
7.5	7.500	7.493	-0.007	41156	0.0 %	41156					0
10.0	10.000	10.046	0.046	55241	0.0 %	55241					0

LB121007 INSTRUMENT ID : CV1

Method: 7471B		Operator: Admin		Date of Analysis: 11 Jul 2022 11:53:03				
Sample ID	Extended ID	µ Abs.	Conc.	Std Conc/Method	Units	Date	Type	Type
0.0 - 1	50	-18	-	0.00007471B	PPB	11 Jul 2022 11:56:05	S	Std
0.2 - 1	50.2	1051	-	0.20007471B	PPB	11 Jul 2022 11:58:22	S	Std
2.5 - 1	52.5	13466	-	2.50007471B	PPB	11 Jul 2022 12:00:39	S	Std
5.0 - 1	55	26998	-	5.00007471B	PPB	11 Jul 2022 12:02:56	S	Std
7.5 - 1	57.5	41156	-	7.50007471B	PPB	11 Jul 2022 12:05:13	S	Std
10.0 - 1	59	55241	-	10.00007471B	PPB	11 Jul 2022 12:07:30	S	Std
ICV33 - 1	ICV33	21944	4.0118	-7471B	PPB	11 Jul 2022 12:10:16	U	SMPL
ICB33 - 1	ICB33	-92	0.0186	-7471B	PPB	11 Jul 2022 12:12:31	U	SMPL
CCV13 - 1	CCV13	27003	4.9286	-7471B	PPB	11 Jul 2022 12:14:47	U	SMPL
CCB13 - 1	CCB13	-93	0.0184	-7471B	PPB	11 Jul 2022 12:17:02	U	SMPL
CRA - 1	CRA	1064	0.2281	-7471B	PPB	11 Jul 2022 12:19:18	U	SMPL
HighStd - 1	HighStd	54969	9.9965	-7471B	PPB	11 Jul 2022 12:21:33	U	SMPL
ChkStd - 1	ChkStd	38187	6.9553	-7471B	PPB	11 Jul 2022 12:23:49	U	SMPL
PB146130BL - 1	PBS	-141	0.0097	-7471B	PPB	11 Jul 2022 12:26:06	U	SMPL
PB146130BS - 1	LCSS	20133	3.6837	-7471B	PPB	11 Jul 2022 12:28:24	U	SMPL
N3214-01 - 1	LOD-MDL-SOIL-01-QT2-2022	109	0.0550	-7471B	PPB	11 Jul 2022 12:30:40	U	SMPL
N3214-02 - 1	LOQ-SOIL-02-QT2-2022	945	0.2065	-7471B	PPB	11 Jul 2022 12:32:57	U	SMPL
N3619-01 - 1	31653	113229	20.5541	-7471B	PPB	11 Jul 2022 12:35:14	U	SMPL
N3619-02 - 1	52322-A-B-C	7054	1.3135	-7471B	PPB	11 Jul 2022 12:37:31	U	SMPL
N3619-03 - 1	31440	293	0.0883	-7471B	PPB	11 Jul 2022 12:39:50	U	SMPL
CCV14 - 1	CCV14	27212	4.9665	-7471B	PPB	11 Jul 2022 12:42:05	U	SMPL
CCB14 - 1	CCB14	-94	0.0182	-7471B	PPB	11 Jul 2022 12:44:20	U	SMPL
N3641-01 - 1	CL-01-070822	3067	0.5910	-7471B	PPB	11 Jul 2022 12:46:37	U	SMPL
N3641-01DUP - 1	CL-01-070822DUP	3364	0.6448	-7471B	PPB	11 Jul 2022 12:48:52	U	SMPL
N3641-01MS - 1	CL-01-070822MS	25557	4.6666	-7471B	PPB	11 Jul 2022 12:51:07	U	SMPL
N3641-01MSD - 1	CL-01-070822MSD	23927	4.3712	-7471B	PPB	11 Jul 2022 12:53:23	U	SMPL
N3643-01 - 1	TR-04-070822	1277	0.2667	-7471B	PPB	11 Jul 2022 12:55:40	U	SMPL
N3644-01 - 1	OR-03-070822	463	0.1191	-7471B	PPB	11 Jul 2022 12:57:58	U	SMPL
N3645-01 - 1	NB-08-070822	2013	0.4000	-7471B	PPB	11 Jul 2022 13:00:14	U	SMPL
N3648-01 - 1	CMR-TP1-0708	4089	0.7762	-7471B	PPB	11 Jul 2022 13:02:31	U	SMPL
N3648-04 - 1	CMR-TP2-0708	10395	1.9190	-7471B	PPB	11 Jul 2022 13:04:48	U	SMPL
N3649-01 - 1	B-2-UC	440	0.1150	-7471B	PPB	11 Jul 2022 13:07:03	U	SMPL
CCV15 - 1	CCV15	27158	4.9567	-7471B	PPB	11 Jul 2022 13:09:19	U	SMPL
CCB15 - 1	CCB15	-100	0.0171	-7471B	PPB	11 Jul 2022 13:11:34	U	SMPL
N3649-02 - 1	A-3-UC	1738	0.3502	-7471B	PPB	11 Jul 2022 13:13:51	U	SMPL
N3649-03 - 1	A-3-LC	5437	1.0205	-7471B	PPB	11 Jul 2022 13:16:07	U	SMPL
N3649-04 - 1	A-6-UC	295	0.0887	-7471B	PPB	11 Jul 2022 13:18:22	U	SMPL
N3649-05 - 1	A-6-LC	576	0.1396	-7471B	PPB	11 Jul 2022 13:20:38	U	SMPL
N3649-06 - 1	A-2-UC	2826	0.5474	-7471B	PPB	11 Jul 2022 13:22:53	U	SMPL
N3649-07 - 1	A-2-LC	1538	0.3139	-7471B	PPB	11 Jul 2022 13:25:10	U	SMPL
N3641-01LX5 - 1		494	0.1248	-7471B	PPB	11 Jul 2022 13:27:26	U	SMPL
N3641-01A - 1		24989	4.5636	-7471B	PPB	11 Jul 2022 13:29:43	U	SMPL
N3619-01DLX5 - 1		25197	4.6013	-7471B	PPB	11 Jul 2022 13:32:00	U	SMPL
CCV16 - 1	CCV16	26957	4.9203	-7471B	PPB	11 Jul 2022 13:34:16	U	SMPL
CCB16 - 1	CCB16	-89	0.0191	-7471B	PPB	11 Jul 2022 13:36:33	U	SMPL