

LB121010


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

40675



A= 0.0000e+000

B= 2.4866e-004

C= -1.1334e-001

Rho= 0.9999630

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	9/6/11
0.0	0.000	0.024	0.024	551	0.000	551					-
0.2	0.200	0.175	-0.025	1159	0.0 %	1159					-13
2.5	2.500	2.531	0.031	10635	0.0 %	10635					1
5.0	5.000	4.943	-0.057	20336	0.0 %	20336					-1
7.5	7.500	7.526	0.026	30722	0.0 %	30722					0
10.0	10.000	10.001	0.001	40675	0.0 %	40675					0

LB121010

Method: 7470A

Operator: Admin

Date of Analysis: 11 Jul 2022 13:53:27

Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	551	-	0.0000	7470A	PPB	11 Jul 2022 13:54:57	S	Std
0.2 - 1	50.2	1159	-	0.2000	7470A	PPB	11 Jul 2022 13:57:14	S	Std
2.5 - 1	52.5	10635	-	2.5000	7470A	PPB	11 Jul 2022 13:59:31	S	Std
5.0 - 1	55	20336	-	5.0000	7470A	PPB	11 Jul 2022 14:01:48	S	Std
7.5 - 1	57.5	30722	-	7.5000	7470A	PPB	11 Jul 2022 14:04:04	S	Std
10.0 - 1	59.0	40675	-	10.0000	7470A	PPB	11 Jul 2022 14:08:53	S	Std
ICV34 - 1	ICV34	15664	3.7817	-	7470A	PPB	11 Jul 2022 14:11:32	U	SMPL
ICB34 - 1	ICB34	353	-0.0256	-	7470A	PPB	11 Jul 2022 14:13:46	U	SMPL
CCV17 - 1	CCV17	20213	4.9128	-	7470A	PPB	11 Jul 2022 14:16:02	U	SMPL
CCB17 - 1	CCB17	359	-0.0241	-	7470A	PPB	11 Jul 2022 14:18:16	U	SMPL
CRA - 1	CRA	1115	0.1639	-	7470A	PPB	11 Jul 2022 14:20:32	U	SMPL
HighStd - 1	HighStd	40776	10.0260	-	7470A	PPB	11 Jul 2022 14:22:47	U	SMPL
ChkStd - 1	ChkStd	28643	7.0090	-	7470A	PPB	11 Jul 2022 14:25:02	U	SMPL
PB146146BL - 1	PBW	344	-0.0278	-	7470A	PPB	11 Jul 2022 14:27:19	U	SMPL
PB146146BS - 1	LCSW	15019	3.6213	-	7470A	PPB	11 Jul 2022 14:29:36	U	SMPL
N3214-07 - 1	LOD-MDL-WATER-01-QT2-2022	605	0.0371	-	7470A	PPB	11 Jul 2022 14:39:25	U	SMPL
N3214-08 - 1	LOQ-WATER-02-QT2-2022	1177	0.1793	-	7470A	PPB	11 Jul 2022 14:41:41	U	SMPL
N3627-01 - 1	MW-107	470	0.0035	-	7470A	PPB	11 Jul 2022 14:43:58	U	SMPL
N3627-01DUP - 1	MW-107DUP	487	0.0078	-	7470A	PPB	11 Jul 2022 14:46:15	U	SMPL
N3627-01MS - 1	MW-107MS	14496	3.4912	-	7470A	PPB	11 Jul 2022 14:48:30	U	SMPL
CCV18 - 1	CCV18	19956	4.8489	-	7470A	PPB	11 Jul 2022 14:50:45	U	SMPL
CCB18 - 1	CCB18	421	-0.0087	-	7470A	PPB	11 Jul 2022 14:53:01	U	SMPL
N3627-01MSD - 1	MW-107MSD	14864	3.5827	-	7470A	PPB	11 Jul 2022 14:55:18	U	SMPL
N3627-01LX5 - 1		487	0.0078	-	7470A	PPB	11 Jul 2022 14:57:33	U	SMPL
N3627-01A - 1		15824	3.8215	-	7470A	PPB	11 Jul 2022 14:59:49	U	SMPL
CCV19 - 1	CCV19	19984	4.8559	-	7470A	PPB	11 Jul 2022 15:02:05	U	SMPL
CCB19 - 1	CCB19	401	-0.0136	-	7470A	PPB	11 Jul 2022 15:04:22	U	SMPL