



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Guerda

Parameter: Flash Point

Supervisor Review By: mohammad

Run Number: LB121012

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2743

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.1	07/08/2022	13:05
2	N3590-06		1	12	78.00	172.4	172.1	07/08/2022	13:35
3	N3604-04		1	15	100.00	>212.0	>212.0	07/08/2022	13:58
4	N3604-05		1	18	100.00	>212.0	>212.0	07/08/2022	14:20
5	N3619-04		1	17	100.00	>212.0	>212.0	07/08/2022	14:45
6	N3630-01		1	18	100.00	>212.0	>212.0	07/08/2022	15:05
7	N3631-03		1	17	100.00	>212.0	>212.0	07/08/2022	15:25
8	N3631-03DUP		1	17	100.00	>212.0	>212.0	07/08/2022	15:40

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB121012

WorkList Name : FLASHPOINT-07-07-22 WorkList ID : 161185 Department : Wet-Chemistry Date : 07-07-2022 16:24:38

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
07/12/2022	Water	N3590-06	Flash Point	Cool 4 deg C	PSEG03	161		07/05/2022	1010B
07/13/2022	Water	N3604-04	Flash Point	Cool 4 deg C	PSEG03	G11	1969	07/06/2022	1010B
07/13/2022	Water	N3604-05	Flash Point	Cool 4 deg C	PSEG03	G11	1970	07/06/2022	1010B
07/14/2022	Water	N3619-04	Flash Point	Cool 4 deg C	PSEG03	G11	31654	07/07/2022	1010B
07/14/2022	Water	N3630-01	Flash Point	Cool 4 deg C	PSEG03	G11	33747-57	07/07/2022	1010B
07/14/2022	Water	N3631-03	Flash Point	Cool 4 deg C	PSEG03	G11	33744	07/07/2022	1010B

Date/Time 07/08/22 12:40
 Raw Sample Received by: SKC
 Raw Sample Relinquished by: KS(WC)

Date/Time 07/08/22 15:45
 Raw Sample Received by: KS(WC)
 Raw Sample Relinquished by: SK(WC)