



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Guerda

Parameter: Flash Point

Supervisor Review By: mohammad

Run Number: LB121013

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2743

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.3	07/11/2022	13:10
2	N3602-03		1	15	100.00	>212.0	>212.0	07/11/2022	13:35
3	N3602-03DUP		1	15	100.00	>212.0	>212.0	07/11/2022	13:58
4	N3602-05		1	18	100.00	>212.0	>212.0	07/11/2022	14:20
5	N3602-07		1	17	100.00	>212.0	>212.0	07/11/2022	14:45
6	N3602-09		1	15	100.00	>212.0	>212.0	07/11/2022	15:10

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB121013

WorkList Name : FLASHPOINT-07-11-22 WorkList ID : 161256 Department : Wet-Chemistry Date : 07-11-2022 12:38:59

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
07/13/2022	Solid	N3602-03	Flash Point	Cool 4 deg C	PSEG03	G11	31530	07/06/2022	1010B
07/13/2022	Solid	N3602-05	Flash Point	Cool 4 deg C	PSEG03	G11	21245	07/06/2022	1010B
07/13/2022	Solid	N3602-07	Flash Point	Cool 4 deg C	PSEG03	G11	21244	07/06/2022	1010B
07/13/2022	Solid	N3602-09	Flash Point	Cool 4 deg C	PSEG03	G11	21247	07/06/2022	1010B

Date/Time 07/11/22 12:45
 Raw Sample Received by: KS (WC)
 Raw Sample Relinquished by: KS (WC)

Date/Time 07/11/22 15:15
 Raw Sample Received by: KS (WC)
 Raw Sample Relinquished by: KS (WC)