

LB121016


SFAM01.1

INSTRUMENT ID: CV2

Linear

 μ Abs.:

42911



A= 0.0000e+000

B= 2.3543e-004 *sign*C= -8.3801e-002 *y=10/accept*

Rho= 0.9999332

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	<i>o/oD</i>
0.00	0.000	-0.058	-0.058	108	0.000	108					<i>11</i>
0.05	0.050					185	0				<i>20</i>
0.20	0.200	0.239	0.039	1371	0.0 %	1371					<i>2</i>
2.50	2.500	2.561	0.061	11234	0.0 %	11234					<i>1</i>
5.00	5.000	4.962	-0.038	21434	0.0 %	21434					<i>0</i>
7.50	7.500	7.478	-0.022	32118	0.0 %	32118					<i>0</i>
10.0	10.000	10.019	0.019	42911	0.0 %	42911					<i>0</i>

LB121016

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	108	-		0 SFAM01.1	PPB	7/11/2022 15:44	Std
	0.2 S01	1371	-		0.2 SFAM01.1	PPB	7/11/2022 15:47	Std
	2.5 S02	11234	-		2.5 SFAM01.1	PPB	7/11/2022 15:49	Std
	5 S03	21434	-		5 SFAM01.1	PPB	7/11/2022 15:51	Std
	7.5 S04	32118	-		7.5 SFAM01.1	PPB	7/11/2022 15:54	Std
	10 S05	42911	-		10 SFAM01.1	PPB	7/11/2022 15:57	Std
ICV035	ICV035	17533	4.0439 -		SFAM01.1	PPB	7/11/2022 16:00	SMPL
ICB035	ICB035	62	-0.0692 -		SFAM01.1	PPB	7/11/2022 16:03	SMPL
CCV020	CCV020	21601	5.0016 -		SFAM01.1	PPB	7/11/2022 16:05	SMPL
CCB020	CCB020	56	-0.0706 -		SFAM01.1	PPB	7/11/2022 16:07	SMPL
PB146143BL	PBW143	-24	-0.0895 -		SFAM01.1	PPB	7/11/2022 16:09	SMPL
PB146109TB	LEB143	29	-0.077 -		SFAM01.1	PPB	7/11/2022 16:12	SMPL
N3628-01	MC0ZB8	42	-0.0739 -		SFAM01.1	PPB	7/11/2022 16:14	SMPL
N3628-02	MC0ZB9	63	-0.069 -		SFAM01.1	PPB	7/11/2022 16:16	SMPL
N3628-03	MC0ZB9D	83	-0.0643 -		SFAM01.1	PPB	7/11/2022 16:18	SMPL
N3628-04	MC0ZB9S	4831	1.0535 -		SFAM01.1	PPB	7/11/2022 16:21	SMPL
CCV021	CCV021	21517	4.9818 -		SFAM01.1	PPB	7/11/2022 16:23	SMPL
CCB021	CCB021	58	-0.0701 -		SFAM01.1	PPB	7/11/2022 16:25	SMPL