



Analytical Summary Report

Analysis Method: 9040C
Parameter: pH
Run Number: LB121028

Analyst By : JIGNESH

Supervisor Review By : ketankumar

Slope : 98.8

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2905
BUFFER PH 7.00 GREEN 1PINT PK6	W2906
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	07/12/2022	11:40
2	CAL2	1	Water	NA	NA	20.7	7.00	07/12/2022	11:41
3	CAL3	1	Water	NA	NA	20.7	10.03	07/12/2022	11:42
4	ICV	1	Water	NA	NA	20.7	6.98	07/12/2022	11:45
5	CCV1	1	Water	NA	NA	20.7	2.01	07/12/2022	11:46
6	N3672-02	1	Water	NA	NA	20.1	6.51	07/12/2022	11:50
7	N3672-02DUP	1	Water	NA	NA	20.2	6.52	07/12/2022	11:51
8	N3673-01	1	Water	NA	NA	22.7	5.93	07/12/2022	12:00
9	N3675-01	1	Water	NA	NA	23.1	5.12	07/12/2022	12:05
10	N3675-02	1	Water	NA	NA	23.2	5.22	07/12/2022	12:09
11	CCV2	1	Water	NA	NA	20.8	12.03	07/12/2022	12:10

WORKLIST(Hardcopy Internal Chain)

WB 121028

WorkList Name : PH N3673 WorkList ID : 161284 Department : Wet-Chemistry Date : 07-12-2022 11:34:09

Due Date	Matrix	Sample	Test	Raw Sample			Collect Date	Method
				Preservative	Customer	Storage Location		
07/11/2022	Water	N3672-02	pH ✓	Cool 4 deg C	TETR06	N11	TT-048-IDWGW-20220711	07/11/2022 9040C
07/07/2022	Water	N3673-01	pH ✗	Cool 4 deg C	STER11	N11	POOL-WATER	07/07/2022 9040C
07/08/2022	Water	N3675-01	pH ✗	Cool 4 deg C	STER16	N11	MAIN-CELL	07/08/2022 9040C
07/08/2022	Water	N3675-02	pH ✗	Cool 4 deg C	STER16	N11	EXCELL	07/08/2022 9040C

Date/Time 07/12/2022 11:35
 Raw Sample Received by: 706011
 Raw Sample Relinquished by: 126011

Date/Time 07/12/2022 13:00
 Raw Sample Received by: 126011
 Raw Sample Relinquished by: 706020