

LB121029


7471B

INSTRUMENT ID: CV 1

Linear

 μ Abs.:

53906



A= 0.0000e+000

B= 1.8764e-004

C= 5.6183e-003

Rho= 0.9997548

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/0 D
0.0	0.000	0.087	0.087	436	0.000	436					—
0.2	0.200	0.201	0.001	1043	0.0 %	1043					—
2.5	2.500	2.472	-0.028	13145	0.0 %	13145					—
5.0	5.000	4.926	-0.074	26223	0.0 %	26223					—
7.5	7.500	7.392	-0.108	39365	0.0 %	39365					—
10.0	10.000	10.121	0.121	53906	0.0 %	53906					—

Method: 7471B Operator: Admin Date of Analysis: 12 Jul 2022 12:48:00

Date of Analysis: 12 Jul 2022 12:48:00

Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	436	-	0.0000	7471B	PPB	12 Jul 2022 13:37:54	S	Std
0.2 - 1	50.2	1043	-	0.2000	7471B	PPB	12 Jul 2022 13:40:11	S	Std
2.5 - 1	52.5	13145	-	2.5000	7471B	PPB	12 Jul 2022 13:42:29	S	Std
5.0 - 1	55	26223	-	5.0000	7471B	PPB	12 Jul 2022 13:44:46	S	Std
7.5 - 1	57.5	39365	-	7.5000	7471B	PPB	12 Jul 2022 13:47:02	S	Std
10.0 - 1	510	53906	-	10.0000	7471B	PPB	12 Jul 2022 13:49:20	S	Std
ICV36 - 1	ICV36	22110	4.1544	-	7471B	PPB	12 Jul 2022 13:52:28	U	SMPL
ICB36 - 1	ICB36	-2	0.0052	-	7471B	PPB	12 Jul 2022 13:54:42	U	SMPL
CCV22 - 1	CCV22	26405	4.9603	-	7471B	PPB	12 Jul 2022 13:56:58	U	SMPL
CCB22 - 1	CCB22	0	0.0056	-	7471B	PPB	12 Jul 2022 13:59:12	U	SMPL
CRA - 1	CRA	1036	0.2000	-	7471B	PPB	12 Jul 2022 14:01:29	U	SMPL
HighStd - 1	HighStd	54008	10.1398	-	7471B	PPB	12 Jul 2022 14:03:44	U	SMPL
ChkStd - 1	ChkStd	37847	7.1073	-	7471B	PPB	12 Jul 2022 14:05:59	U	SMPL
PB146166BL - 1	PBS	-65	-0.0066	-	7471B	PPB	12 Jul 2022 14:08:17	U	SMPL
PB146166BS - 1	LCSS	21056	3.9566	-	7471B	PPB	12 Jul 2022 14:10:35	U	SMPL
N3214-01 - 1	LOD-MDL-SOIL-01-QT2-2022	243	0.0512	-	7471B	PPB	12 Jul 2022 14:12:51	U	SMPL
N3214-01RE - 1	LOD-MDL-SOIL-01-QT2-2022RE	700	0.1370	-	7471B	PPB	12 Jul 2022 14:15:09	U	SMPL
N3214-03 - 1	MDL-SOIL-03-QT2-2022	307	0.0632	-	7471B	PPB	12 Jul 2022 14:17:25	U	SMPL
N3662-01 - 1	20220495A-SIM/C1A-COMPOSITE	84195	15.8042	-	7471B	PPB	12 Jul 2022 14:19:42	U	SMPL
N3662-02 - 1	20220495B-SIM/C1B-COMPOSITE	78002	14.6421	-	7471B	PPB	12 Jul 2022 14:21:57	U	SMPL
CCV23 - 1	CCV23	26400	4.9594	-	7471B	PPB	12 Jul 2022 14:24:14	U	SMPL
CCB23 - 1	CCB23	-32	-0.0004	-	7471B	PPB	12 Jul 2022 14:26:32	U	SMPL
N3662-03 - 1	20220496B-SIM/C2B-COMPOSITE	2140	0.4072	-	7471B	PPB	12 Jul 2022 14:28:49	U	SMPL
N3662-04 - 1	20220497B-SIMC3B-COMPOSITE	248002	46.5414	-	7471B	PPB	12 Jul 2022 14:31:04	U	SMPL
N3662-05 - 1	20220498B-SIM/C4B-COMPOSITE	198568	37.2655	-	7471B	PPB	12 Jul 2022 14:33:19	U	SMPL
N3662-05DUP - 1	20220498B-SIM/C4B-COMPOSITEDUP	191111	35.8662	-	7471B	PPB	12 Jul 2022 14:35:42	U	SMPL
N3662-05MS - 1	20220498B-SIM/C4B-COMPOSITEDMS	220866	41.4496	-	7471B	PPB	12 Jul 2022 14:38:04	U	SMPL
N3662-05MSD - 1	20220498B-SIM/C4B-COMPOSITEDMSD	222454	41.7475	-	7471B	PPB	12 Jul 2022 14:40:26	U	SMPL
N3663-01 - 1	VB16181	53213	9.9907	-	7471B	PPB	12 Jul 2022 14:42:49	U	SMPL
N3662-05LX5 - 1		43041	8.0820	-	7471B	PPB	12 Jul 2022 14:45:12	U	SMPL
N3662-05A - 1		224000	42.0376	-	7471B	PPB	12 Jul 2022 14:47:32	U	SMPL
CCV24 - 1	CCV24	25865	4.8590	-	7471B	PPB	12 Jul 2022 14:49:49	U	SMPL
CCB24 - 1	CCB24	-72	-0.0079	-	7471B	PPB	12 Jul 2022 14:52:11	U	SMPL
N3662-01DLX5 - 1		17571	3.3027	-	7471B	PPB	12 Jul 2022 14:54:28	U	SMPL
N3662-02DLX5 - 1		16564	3.1137	-	7471B	PPB	12 Jul 2022 14:56:44	U	SMPL
N3662-04DLX10 - 1		27712	5.2056	-	7471B	PPB	12 Jul 2022 14:59:00	U	SMPL
N3662-05DLX10 - 1		21810	4.0981	-	7471B	PPB	12 Jul 2022 15:01:16	U	SMPL
N3662-05DUPDLX10 - 1		19979	3.7545	-	7471B	PPB	12 Jul 2022 15:03:33	U	SMPL
N3662-05MSDLX10 - 1		24246	4.5552	-	7471B	PPB	12 Jul 2022 15:08:42	U	SMPL
N3662-05MSDDLX10 - 1		22850	4.2933	-	7471B	PPB	12 Jul 2022 15:10:58	U	SMPL
N3662-05LDLX50 - 1		3735	0.7065	-	7471B	PPB	12 Jul 2022 15:13:15	U	SMPL
N3662-05ADLX10 - 1		23955	4.5006	-	7471B	PPB	12 Jul 2022 15:15:34	U	SMPL
CCV25 - 1	CCV25	26279	4.9367	-	7471B	PPB	12 Jul 2022 15:17:50	U	SMPL
CCB25 - 1	CCB25	-18	0.0022	-	7471B	PPB	12 Jul 2022 15:20:07	U	SMPL