



Analytical Summary Report

Analysis Method: 9040C

Parameter: Corrosivity

Run Number: LB121039

Analyst By : JIGNESH

Supervisor Review By : mohammad

Slope : 99.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2905
BUFFER PH 7.00 GREEN 1PINT PK6	W2906
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.8	4.01	07/13/2022	09:00
2	CAL2	1	Water	NA	NA	20.8	7.00	07/13/2022	09:01
3	CAL3	1	Water	NA	NA	20.08	10.02	07/13/2022	09:02
4	ICV	1	Water	NA	NA	20.7	6.98	07/13/2022	09:05
5	CCV1	1	Water	NA	NA	20.7	2.01	07/13/2022	09:06
6	N3682-05	1	Water	NA	NA	25.6	8.33	07/13/2022	09:25
7	N3682-05DUP	1	Water	NA	NA	25.7	8.34	07/13/2022	09:26
8	N3682-06	1	Water	NA	NA	25.3	8.87	07/13/2022	09:30
9	CCV2	1	Water	NA	NA	20.5	12.01	07/13/2022	09:33

VB 121039

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity n3682 w WorkList ID : 161310 Department : Wet-Chemistry Date : 07-13-2022 08:47:09

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample		Collect Date	Method
						Storage	Location		
07/12/2022	Water	N3682-05	Corrosivity	Cool 4 deg C	PSEG03	G12	17172	07/12/2022	9040C
07/12/2022	Water	N3682-06	Corrosivity	Cool 4 deg C	PSEG03	G12	17166	07/12/2022	9040C

Date/Time 07-13-22 08:50
Raw Sample Received by: JP (WC)
Raw Sample Relinquished by: RS CEF (106)

Date/Time 07-13-22 11:00
Raw Sample Received by: RS CEF (106)
Raw Sample Relinquished by: JP (WC)