

LB121040


7470A

INSTRUMENT ID: CVZ

Linear

 μ Abs.:

39560



A= 0.0000e+000

B= 2.5255e-004

C= -6.9478e-002

Rho= 0.9999107

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	good
0.0	0.000	-0.018	-0.018	203	0.000	203					-11
0.2	0.200	0.179	-0.021	984	0.0 %	984					0
2.5	2.500	2.509	0.009	10210	0.0 %	10210					1
5.0	5.000	5.025	0.025	20172	0.0 %	20172					1
7.5	7.500	7.583	0.083	30302	0.0 %	30302					1
10.0	10.000	9.922	-0.078	39560	0.0 %	39560					-1

LB121040 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 13 Jul 2022 10:32:22				
Sample ID	Extended ID	μ Abs.	Conc.	Std CondMethod	Units	Date	Type	Type
0.0 - 1	50	203	-	0.0007470A	PPB	13 Jul 2022 10:36:45	S	Std
0.2 - 1	50.2	984	-	0.20007470A	PPB	13 Jul 2022 10:39:03	S	Std
2.5 - 1	52.1	10210	-	2.50007470A	PPB	13 Jul 2022 10:44:37	S	Std
5.0 - 1	55	20172	-	5.00007470A	PPB	13 Jul 2022 10:46:54	S	Std
7.5 - 1	59.5	30302	-	7.50007470A	PPB	13 Jul 2022 10:54:20	S	Std
10.0 - 1	57.5	39560	-	10.00007470A	PPB	13 Jul 2022 11:01:41	S	Std
ICV37 - 1	ICV37	17002	4.2245	-7470A	PPB	13 Jul 2022 11:04:24	U	SMPL
ICB37 - 1	ICB37	-34	-0.0781	-7470A	PPB	13 Jul 2022 11:06:39	U	SMPL
CCV26 - 1	CCV26	20574	5.1266	-7470A	PPB	13 Jul 2022 11:08:54	U	SMPL
CCB26 - 1	CCB26	-32	-0.0776	-7470A	PPB	13 Jul 2022 11:11:09	U	SMPL
CRA - 1	CRA	1135	0.2172	-7470A	PPB	13 Jul 2022 11:18:37	U	SMPL
HighStd - 1	HighStd	38950	9.7675	-7470A	PPB	13 Jul 2022 11:20:51	U	SMPL
ChkStd - 1	ChkStd	28517	7.1326	-7470A	PPB	13 Jul 2022 11:23:07	U	SMPL
PB146161BL - 1	PBW	-59	-0.0844	-7470A	PPB	13 Jul 2022 11:25:23	U	SMPL
PB146161BS - 1	LCSW	14828	3.6754	-7470A	PPB	13 Jul 2022 11:27:40	U	SMPL
N3516-04 - 1	R-1	-28	-0.0765	-7470A	PPB	13 Jul 2022 11:29:56	U	SMPL
N3516-04DUP - 1	R-1DUP	12	-0.0664	-7470A	PPB	13 Jul 2022 11:32:12	U	SMPL
N3516-04MS - 1	R-1MS	14274	3.5355	-7470A	PPB	13 Jul 2022 11:34:29	U	SMPL
N3516-04MSD - 1	R-1MSD	16651	4.1358	-7470A	PPB	13 Jul 2022 11:51:29	U	SMPL
PB146163BL - 1	PBW	-33	-0.0778	-7470A	PPB	13 Jul 2022 11:53:44	U	SMPL
CCV27 - 1	CCV27	19558	4.8700	-7470A	PPB	13 Jul 2022 11:55:59	U	SMPL
CCB27 - 1	CCB27	-31	-0.0773	-7470A	PPB	13 Jul 2022 11:58:15	U	SMPL
PB146163BS - 1	LCSW	13890	3.4385	-7470A	PPB	13 Jul 2022 12:00:31	U	SMPL
N3648-02 - 1	CMR-TP1-0708	-41	-0.0798	-7470A	PPB	13 Jul 2022 12:02:46	U	SMPL
N3648-05 - 1	CMR-TP2-0708	7	-0.0677	-7470A	PPB	13 Jul 2022 12:05:01	U	SMPL
N3663-02 - 1	VB16181	13	-0.0662	-7470A	PPB	13 Jul 2022 12:07:17	U	SMPL
N3663-02DUP - 1	VB16181DUP	7	-0.0677	-7470A	PPB	13 Jul 2022 12:09:33	U	SMPL
N3663-02MS - 1	VB16181MS	15170	3.7618	-7470A	PPB	13 Jul 2022 12:11:49	U	SMPL
N3663-02MSD - 1	VB16181MSD	13755	3.4044	-7470A	PPB	13 Jul 2022 12:14:05	U	SMPL
PB146167BL - 1	PBW	-31	-0.0773	-7470A	PPB	13 Jul 2022 12:16:22	U	SMPL
PB146167BS - 1	LCSW	17096	4.2482	-7470A	PPB	13 Jul 2022 12:18:38	U	SMPL
N3214-07 - 1	LOD-MDL-WATER-01-QT2-2022	447	0.0434	-7470A	PPB	13 Jul 2022 12:30:49	U	SMPL
CCV28 - 1	CCV28	20373	5.0758	-7470A	PPB	13 Jul 2022 12:33:04	U	SMPL
CCB28 - 1	CCB28	-34	-0.0781	-7470A	PPB	13 Jul 2022 12:35:20	U	SMPL
N3214-07RE - 1	LOD-MDL-WATER-01-QT2-2022RE	849	0.1449	-7470A	PPB	13 Jul 2022 12:37:36	U	SMPL
N3214-09 - 1	MDL-WATER-03-QT2-2022	447	0.0434	-7470A	PPB	13 Jul 2022 12:47:33	U	SMPL
PB146161TB - 1	PB146144TB	1	-0.0692	-7470A	PPB	13 Jul 2022 12:49:48	U	SMPL
PB146145TB - 1	PB146145TB	14	-0.0659	-7470A	PPB	13 Jul 2022 12:52:04	U	SMPL
N3516-04LX5 - 1	PB146144TB	-42	-0.0801	-7470A	PPB	13 Jul 2022 12:54:20	U	SMPL
N3516-04A - 1		15206	3.7709	-7470A	PPB	13 Jul 2022 12:56:36	U	SMPL
N3663-02LX5 - 1		-94	-0.0932	-7470A	PPB	13 Jul 2022 12:58:53	U	SMPL
N3663-02A - 1		17111	4.2520	-7470A	PPB	13 Jul 2022 13:01:09	U	SMPL
CCV29 - 1	CCV29	19656	4.8947	-7470A	PPB	13 Jul 2022 13:03:26	U	SMPL
CCB29 - 1	CCB29	-17	-0.0738	-7470A	PPB	13 Jul 2022 13:05:42	U	SMPL