

L3121050


7471B

INSTRUMENT ID: CV1

Linear

μ Abs.:

53164



A= 0.0000e+000

B= 1.8694e-004

C= 4.5759e-002

Rho= 0.9999106

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.041	0.041	-24	0.000	-24					-6
0.2	0.200	0.213	0.013	893	0.0 %	893					-2
2.5	2.500	2.444	-0.056	12831	0.0 %	12831					-1
5.0	5.000	4.941	-0.059	26183	0.0 %	26183					1
7.5	7.500	7.577	0.077	40284	0.0 %	40284					0
10.0	10.000	9.984	-0.016	53164	0.0 %	53164					

LB121050 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 13 Jul 2022 14:15:04

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-24	-	0.0000	7471B	PPB	13 Jul 2022 14:16:41	S	Std
0.2 - 1	51.2	893	-	0.2000	7471B	PPB	13 Jul 2022 14:18:58	S	Std
2.5 - 1	52.5	12831	-	2.5000	7471B	PPB	13 Jul 2022 14:21:15	S	Std
5.0 - 1	55	26183	-	5.0000	7471B	PPB	13 Jul 2022 14:23:32	S	Std
7.5 - 1	57.5	40284	-	7.5000	7471B	PPB	13 Jul 2022 14:25:48	S	Std
10.0 - 1	510	53164	-	10.0000	7471B	PPB	13 Jul 2022 14:30:39	S	Std
ICV38 - 1	ICV38	21963	4.1516	-7471B	PPB	PPB	13 Jul 2022 14:33:23	U	SMPL
ICB38 - 1	ICB38	-89	0.0291	-7471B	PPB	PPB	13 Jul 2022 14:35:37	U	SMPL
CCV30 - 1	CCV30	26074	4.9201	-7471B	PPB	PPB	13 Jul 2022 14:43:09	U	SMPL
CCB30 - 1	CCB30	-115	0.0243	-7471B	PPB	PPB	13 Jul 2022 14:45:24	U	SMPL
CRA - 1	CRA	845	0.2037	-7471B	PPB	PPB	13 Jul 2022 14:47:40	U	SMPL
HighStd - 1	HighStd	53179	9.9873	-7471B	PPB	PPB	13 Jul 2022 14:49:55	U	SMPL
ChkStd - 1	ChkStd	38318	7.2091	-7471B	PPB	PPB	13 Jul 2022 14:52:11	U	SMPL
PB146204BL - 1	PBS	-151	0.0175	-7471B	PPB	PPB	13 Jul 2022 14:54:28	U	SMPL
PB146204BS - 1	LCSS	19963	3.7777	-7471B	PPB	PPB	13 Jul 2022 14:56:46	U	SMPL
N3687-01 - 1	ETGI-361	9053	1.7382	-7471B	PPB	PPB	13 Jul 2022 14:59:02	U	SMPL
N3687-03 - 1	ETGI-346	27928	5.2667	-7471B	PPB	PPB	13 Jul 2022 15:01:19	U	SMPL
N3688-01 - 1	72-12016	1677	0.3593	-7471B	PPB	PPB	13 Jul 2022 15:03:37	U	SMPL
N3688-01DUP - 1	72-12016DUP	1776	0.3778	-7471B	PPB	PPB	13 Jul 2022 15:05:55	U	SMPL
N3688-01MS - 1	72-12016MS	21749	4.1116	-7471B	PPB	PPB	13 Jul 2022 15:08:10	U	SMPL
CCV31 - 1	CCV31	25400	4.7941	-7471B	PPB	PPB	13 Jul 2022 15:10:26	U	SMPL
CCB31 - 1	CCB31	-80	0.0308	-7471B	PPB	PPB	13 Jul 2022 15:12:43	U	SMPL
N3688-01MSD - 1	72-12016MSD	21149	3.9994	-7471B	PPB	PPB	13 Jul 2022 15:15:00	U	SMPL
N3688-03 - 1	VNJ-240	14248	2.7093	-7471B	PPB	PPB	13 Jul 2022 15:17:15	U	SMPL
N3688-01LX5 - 1		259	0.0942	-7471B	PPB	PPB	13 Jul 2022 15:28:23	U	SMPL
N3688-01A - 1		21308	4.0292	-7471B	PPB	PPB	13 Jul 2022 15:30:38	U	SMPL
CCV32 - 1	CCV32	25981	4.9028	-7471B	PPB	PPB	13 Jul 2022 15:32:54	U	SMPL
CCB32 - 1	CCB32	-91	0.0287	-7471B	PPB	PPB	13 Jul 2022 15:35:12	U	SMPL