

CHEMTECH CONSULTING GROUP INC
284 Sheffield Street, Mountainside, NJ 07092

Reviewed by : SM Instrument ID : Kone

7/14/2022 8:45

Test: Phenolics-

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	1.042	0.0	0.107	
ICB1	0.001	0.0	0.007	
CCV1	1.046	0.0	0.107	
CCB1	0.001	0.0	0.007	
PB146205BL	0.000	0.0	0.007	
PB146205BS	1.059	0.0	0.108	
N3214-01	0.024	0.0	0.010	
N3214-02	0.049	0.0	0.012	
PB146206BL	0.000	0.0	0.007	
PB146206BS	1.053	0.0	0.108	
N3214-07	0.021	0.0	0.009	
N3214-08	0.048	0.0	0.012	
CCV2	1.048	0.0	0.107	
CCB2	-0.003	0.0	0.007	

N	14
Mean	0.385
SD	0.5143
CV%	133.63

Aquakem v. 7.2AQ1

Results from time period:

Wed Jul 13 16:09:40 2022

Wed Jul 13 16:23:41 2022

Sample Id	Sam/Ctr/c#	Test short	Test type	Result	Result unit	Result date and time
0.00PPM	A	Phenolics-	P	0.0062	mg/l	7/13/2022 13:45:13
Phenol-2	A	Phenolics-	P	0.0583	mg/l	7/13/2022 13:45:14
Phenol-2	A	Phenolics-	P	0.1052	mg/l	7/13/2022 13:45:15
Phenol-2	A	Phenolics-	P	0.2466	mg/l	7/13/2022 13:45:16
Phenol-2	A	Phenolics-	P	0.5005	mg/l	7/13/2022 13:45:17
Phenol-2	A	Phenolics-	P	0.9666	mg/l	7/13/2022 13:45:18
Phenol-2	A	Phenolics-	P	2.0165	mg/l	7/13/2022 13:45:19
ICV1	S	Phenolics-	P	1.0419	mg/l	7/13/2022 16:09:41
ICB1	S	Phenolics-	P	0.0006	mg/l	7/13/2022 16:09:42
CCV1	S	Phenolics-	P	1.0459	mg/l	7/13/2022 16:09:44
CCB1	S	Phenolics-	P	0.0006	mg/l	7/13/2022 16:09:46
PB146205BL	S	Phenolics-	P	0.0002	mg/l	7/13/2022 16:09:48
PB146205BS	S	Phenolics-	P	1.0589	mg/l	7/13/2022 16:09:50
N3214-01	S	Phenolics-	P	0.0238	mg/l	7/13/2022 16:19:43
N3214-02	S	Phenolics-	P	0.0491	mg/l	7/13/2022 16:19:44
PB146206BL	S	Phenolics-	P	0	mg/l	7/13/2022 16:19:47
PB146206BS	S	Phenolics-	P	1.0526	mg/l	7/13/2022 16:19:48
N3214-07	S	Phenolics-	P	0.0214	mg/l	7/13/2022 16:19:50
N3214-08	S	Phenolics-	P	0.0482	mg/l	7/13/2022 16:19:52
CCV2	S	Phenolics-	P	1.0479	mg/l	7/13/2022 16:23:40
CCB2	S	Phenolics-	P	-0.0028	mg/l	7/13/2022 16:23:41

CHEMTECH CONSULTING GROUP INC
284 Sheffield Street, Mountainside, NJ 07092

Reviewed by : 12 Instrument ID : Kon

7/13/2022 13:56

Test Phenolics-


Accepted 7/13/2022 13:45

Factor 10.49

Bias 0.007

Coeff. of det. 0.999514

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.00PPM	0.008	0.0062	0.0000	-
2	Phenol-2	0.013	0.0583	0.0500	16.6
3	Phenol-2	0.017	0.1052	0.1000	5.2
4	Phenol-2	0.031	0.2466	0.2500	-1.4
5	Phenol-2	0.055	0.5005	0.5000	0.1
6	Phenol-2	0.099	0.9666	1.0000	-3.3
7	Phenol-2	0.199	2.0165	2.0000	0.6

12 7/13/22