



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB121062

Analyst By : JIGNESH

Supervisor Review By : mohammad

Slope : 99.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2905
BUFFER PH 7.00 GREEN 1PINT PK6	W2906
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.6	4.01	07/14/2022	10:15
2	CAL2	1	Water	NA	NA	20.6	7.00	07/14/2022	10:16
3	CAL3	1	Water	NA	NA	20.6	10.03	07/14/2022	10:17
4	ICV	1	Water	NA	NA	20.7	6.98	07/14/2022	10:19
5	CCV1	1	Water	NA	NA	20.7	2.02	07/14/2022	10:22
6	N3696-01	1	Water	NA	NA	22.9	4.82	07/14/2022	10:25
7	N3697-01	1	Water	NA	NA	20.5	13.34	07/14/2022	10:30
8	N3697-02	1	Water	NA	NA	20.6	8.80	07/14/2022	10:35
9	N3707-01	1	Water	NA	NA	25.0	6.99	07/14/2022	10:37
10	N3707-01DUP	1	Water	NA	NA	25.1	7.00	07/14/2022	10:38
11	N3711-09	1	Water	NA	NA	20.6	9.05	07/14/2022	10:42
12	CCV2	1	Water	NA	NA	20.7	12.03	07/14/2022	10:45

WORKLIST(Hardcopy Internal Chain)

VB121062

WorkList Name : ph n3707 WorkList ID : 161355 Department : Wet-Chemistry Date : 07-14-2022 10:05:35

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
07/12/2022	Water	N3696-01	pH <i>H</i>	Cool 4 deg C	STER17	N11	JS9600	07/12/2022	9040C
07/13/2022	Water	N3697-01	pH <i>D</i>	Cool 4 deg C	TULL02	G11	PIPELINE-PIGGING-TK2045L	07/13/2022	9040C
07/13/2022	Water	N3697-02	pH <i>A</i>	Cool 4 deg C	TULL02	G11	CASING-NUMBER-5D-WATERING	07/13/2022	9040C
07/13/2022	Water	N3707-01	pH <i>L</i>	Cool 4 deg C	PSEG03	G11	N48967	07/13/2022	9040C
07/13/2022	Water	N3711-09	pH <i>M</i>	Cool 4 deg C	PSEG03	G11	1-4-TOTES-COMP	07/13/2022	9040C

Date/Time 07-14-22 10:10
 Raw Sample Received by: JD WGC
 Raw Sample Relinquished by: CF SM

Date/Time 07-14-22 12:00
 Raw Sample Received by: CF SM
 Raw Sample Relinquished by: JD WGC