



PERCENT SOLID

Supervisor: mohammad

Analyst: JIGNESH

Date: 7/15/2022

OVENTEMP IN Celsius(°C): 107

Time IN: 17:15

In Date: 07/14/2022

Weight Check 1.0g: 1.00

Weight Check 10g: 10.00

OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103

Time OUT: 08:26

Out Date: 07/15/2022

Weight Check 1.0g: 1.00

Weight Check 10g: 10.00

BalanceID: M SC-4

Thermometer ID: %SOLIDS-OVEN

QC:LB121068

| Lab ID   | Client SampleID   | Dish # | Dish Wt(g) (A) | Sample Wt(g) | Dish + Sample Wt(g) (B) | Dish+Dry Sample Wt(g) (C) | % Solid | Comments    |
|----------|-------------------|--------|----------------|--------------|-------------------------|---------------------------|---------|-------------|
| N3711-01 | PSC124019         | 1      | 0.91           | 8.70         | 9.61                    | 7.75                      | 78.6    | GEL MATRIX  |
| N3711-03 | PSC124042         | 2      | 0.93           | 8.93         | 9.86                    | 7.44                      | 72.9    | GEL MATRIX  |
| N3711-05 | PSC124035         | 3      | 0.91           | 9.03         | 9.94                    | 7.78                      | 76.1    |             |
| N3711-07 | PSC124055         | 4      | 0.88           | 8.86         | 9.74                    | 7.12                      | 70.4    |             |
| N3716-01 | COMP-1            | 5      | 0.91           | 8.94         | 9.85                    | 8.72                      | 87.4    |             |
| N3716-02 | COMP-1-EPH-2      | 6      | 0.91           | 8.83         | 9.74                    | 8.94                      | 90.9    |             |
| N3716-03 | VNJ-236           | 7      | 0.93           | 8.61         | 9.54                    | 8.71                      | 90.4    |             |
| N3716-04 | VNJ-236-EPH-2     | 8      | 0.92           | 8.64         | 9.56                    | 8.61                      | 89.0    |             |
| N3716-05 | ETGI-290          | 9      | 0.91           | 8.92         | 9.83                    | 8.98                      | 90.5    |             |
| N3716-06 | ETGI-290-EPH-2    | 10     | 0.91           | 8.86         | 9.77                    | 8.87                      | 89.8    |             |
| N3720-01 | RIDGEFIELD-M&R-SP | 16     | 0.92           | 9.00         | 9.92                    | 9.68                      | 97.3    |             |
| N3720-02 | RIDGEFIELD-M&R-SP | 17     | 0.91           | 8.90         | 9.81                    | 9.49                      | 96.4    |             |
| N3720-03 | PIPE-1            | 18     | 1.00           | 1.00         | 2.00                    | 2.00                      | 100.0   | WIPE SAMPLE |
| N3720-04 | PIPE-2            | 19     | 1.00           | 1.00         | 2.00                    | 2.00                      | 100.0   | WIPE SAMPLE |
| N3720-05 | PIPE-3            | 20     | 1.00           | 1.00         | 2.00                    | 2.00                      | 100.0   | WIPE SAMPLE |
| N3720-06 | PIPE-4            | 21     | 1.00           | 1.00         | 2.00                    | 2.00                      | 100.0   | WIPE SAMPLE |
| N3720-07 | PIPE-5            | 22     | 1.00           | 1.00         | 2.00                    | 2.00                      | 100.0   | WIPE SAMPLE |
| N3720-08 | PIPE-6            | 23     | 1.00           | 1.00         | 2.00                    | 2.00                      | 100.0   | WIPE SAMPLE |
| N3720-09 | PIPE-7            | 24     | 1.00           | 1.00         | 2.00                    | 2.00                      | 100.0   | WIPE SAMPLE |
| N3720-10 | PIPE-8            | 25     | 1.00           | 1.00         | 2.00                    | 2.00                      | 100.0   | WIPE SAMPLE |
| N3720-11 | PIPE-9            | 26     | 1.00           | 1.00         | 2.00                    | 2.00                      | 100.0   | WIPE SAMPLE |
| N3720-12 | PIPE-10           | 27     | 1.00           | 1.00         | 2.00                    | 2.00                      | 100.0   | WIPE SAMPLE |
| N3724-01 | OE-1-071422       | 11     | 0.92           | 8.96         | 9.88                    | 9.17                      | 92.1    |             |
| N3724-02 | OE-1-071422-EPH-2 | 12     | 0.91           | 8.86         | 9.77                    | 9.28                      | 94.5    |             |
| N3727-01 | LOT-071-12        | 15     | 1.00           | 1.00         | 2.00                    | 2.00                      | 100.0   |             |
| N3729-01 | K328-B2           | 28     | 0.92           | 8.76         | 9.68                    | 9.31                      | 95.8    |             |
| N3729-02 | K328-S1           | 29     | 0.96           | 8.98         | 9.94                    | 9.46                      | 94.7    |             |
| N3729-03 | K328-DUP1         | 30     | 0.91           | 8.72         | 9.63                    | 9.23                      | 95.4    |             |



PERCENT SOLID

Supervisor: mohammad

Analyst: JIGNESH

Date: 7/15/2022

OVENTEMP IN Celsius(°C): 107

Time IN: 17:15

In Date: 07/14/2022

Weight Check 1.0g: 1.00

Weight Check 10g: 10.00

OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103

Time OUT: 08:26

Out Date: 07/15/2022

Weight Check 1.0g: 1.00

Weight Check 10g: 10.00

BalanceID: M SC-4

Thermometer ID: %SOLIDS-OVEN

QC:LB121068

| Lab ID   | Client SampleID | Dish # | Dish Wt(g) (A) | Sample Wt(g) | Dish + Sample Wt(g) (B) | Dish+Dry Sample Wt(g) (C) | % Solid | Comments |
|----------|-----------------|--------|----------------|--------------|-------------------------|---------------------------|---------|----------|
| N3731-01 | TP-1            | 13     | 0.88           | 8.78         | 9.66                    | 9.52                      | 98.4    |          |
| N3731-03 | TP-2            | 14     | 0.98           | 8.81         | 9.79                    | 9.35                      | 95.0    |          |

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$