

LB121070


7470A

INSTRUMENT ID: CV2

Linear

 μ Abs.:

39229



A= 0.0000e+000

B= 2.5555e-004

C= -1.7435e-003

Rho= 0.9998038

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.012	0.012	53	0.000	53					-12
0.2	0.200	0.176	-0.024	694	0.0 %	694					-1
2.5	2.500	2.470	-0.030	9673	0.0 %	9673					3
5.0	5.000	5.131	0.131	20087	0.0 %	20087					-1
7.5	7.500	7.388	-0.112	28916	0.0 %	28916					0
10.0	10.000	10.023	0.023	39229	0.0 %	39229					

LB121070 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 14 Jul 2022 11:57:48					
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	53	-	0.0000	7470A	PPB	14 Jul 2022 12:09:17	S	Std
0.2 - 1	50.2	694	-	0.2000	7470A	PPB	14 Jul 2022 12:11:34	S	Std
2.5 - 1	52.5	9673	-	2.5000	7470A	PPB	14 Jul 2022 12:13:51	S	Std
5.0 - 1	55	20087	-	5.0000	7470A	PPB	14 Jul 2022 12:16:08	S	Std
7.5 - 1	57.5	28916	-	7.5000	7470A	PPB	14 Jul 2022 12:18:25	S	Std
10.0 - 1	59.0	39229	-	10.0000	7470A	PPB	14 Jul 2022 12:20:41	S	Std
ICV39 - 1	ICV39	15211	3.8854	-	7470A	PPB	14 Jul 2022 12:23:23	U	SMPL
ICB39 - 1	ICB39	-94	-0.0258	-	7470A	PPB	14 Jul 2022 12:25:38	U	SMPL
CCV33 - 1	CCV33	19873	5.0768	-	7470A	PPB	14 Jul 2022 12:27:53	U	SMPL
CCB33 - 1	CCB33	-56	-0.0161	-	7470A	PPB	14 Jul 2022 12:30:08	U	SMPL
CRA - 1	CRA	659	0.1667	-	7470A	PPB	14 Jul 2022 12:32:24	U	SMPL
HighStd - 1	HighStd	38738	9.8977	-	7470A	PPB	14 Jul 2022 12:34:39	U	SMPL
ChkStd - 1	ChkStd	26083	6.6638	-	7470A	PPB	14 Jul 2022 12:36:54	U	SMPL
PB146226BL - 1	PBW	-84	-0.0232	-	7470A	PPB	14 Jul 2022 12:39:12	U	SMPL
PB146226BS - 1	LCSW	14557	3.7183	-	7470A	PPB	14 Jul 2022 12:41:28	U	SMPL
N3672-02 - 1	TT-048-IDWGW-20220711	-9	-0.0040	-	7470A	PPB	14 Jul 2022 12:43:44	U	SMPL
N3672-02DUP - 1	TT-048-IDWGW-20220711DUP	45	0.0098	-	7470A	PPB	14 Jul 2022 12:46:01	U	SMPL
N3672-02MS - 1	TT-048-IDWGW-20220711MS	14275	3.6462	-	7470A	PPB	14 Jul 2022 12:53:05	U	SMPL
N3672-02MSD - 1	TT-048-IDWGW-20220711MSD	14769	3.7725	-	7470A	PPB	14 Jul 2022 12:55:22	U	SMPL
N3707-01 - 1	N48967	423	0.1064	-	7470A	PPB	14 Jul 2022 12:57:37	U	SMPL
CCV34 - 1	CCV34	20350	5.1987	-	7470A	PPB	14 Jul 2022 12:59:52	U	SMPL
CCB34 - 1	CCB34	-85	-0.0235	-	7470A	PPB	14 Jul 2022 13:02:08	U	SMPL
N3711-09 - 1	1-4-TOTES-COMP	14054	3.5897	-	7470A	PPB	14 Jul 2022 13:04:23	U	SMPL
N3672-02LX5 - 1		-137	-0.0368	-	7470A	PPB	14 Jul 2022 13:06:39	U	SMPL
N3672-02A - 1		15696	4.0094	-	7470A	PPB	14 Jul 2022 13:08:55	U	SMPL
CCV35 - 1	CCV35	20019	5.1141	-	7470A	PPB	14 Jul 2022 13:11:10	U	SMPL
CCB35 - 1	CCB35	-98	-0.0268	-	7470A	PPB	14 Jul 2022 13:13:26	U	SMPL