


LB121080

245.1

 μ Abs.:

37321



Linear

A= 0.0000e+000

B= 2.6844e-004

C= -7.7912e-002

Rho= 0.9998149

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/00
0.0	0.000	-0.051	-0.051	102	0.000	102					10
0.05	0.050	0.055	0.005	495	0.0 %	495	0				30
0.2	0.200	0.260	0.060	1260	0.0 %	1260					4
2.5	2.500	2.397	-0.103	9220	0.0 %	9220					2
5.0	5.000	5.121	0.121	19368	0.0 %	19368					0
7.5	7.500	7.526	0.026	28326	0.0 %	28326					0
10.0	10.000	9.941	-0.059	37321	0.0 %	37321					-1

LB121080 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 14 Jul 2022 15:10:06

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	102	-	0.000	245.1	PPB	14 Jul 2022 15:35:41	S	Std
0.05 - 1	50.05	495	-	0.050	245.1	PPB	14 Jul 2022 15:42:56	S	Std
0.2 - 1	50.2	1260	-	0.200	245.1	PPB	14 Jul 2022 15:45:12	S	Std
2.5 - 1	52.5	9220	-	2.500	245.1	PPB	14 Jul 2022 15:47:29	S	Std
5.0 - 1	55	19368	-	5.000	245.1	PPB	14 Jul 2022 15:49:45	S	Std
7.5 - 1	57.5	28326	-	7.500	245.1	PPB	14 Jul 2022 15:52:01	S	Std
10.0 - 1	510	37321	-	10.000	245.1	PPB	14 Jul 2022 15:54:17	S	Std
ICV40 - 1	ICV40	14985	3.9447	-	245.1	PPB	14 Jul 2022 15:57:18	U	SMPL
ICB40 - 1	ICB40	13	-0.0744	-	245.1	PPB	14 Jul 2022 15:59:33	U	SMPL
CCV36 - 1	CCV36	19476	5.1503	-	245.1	PPB	14 Jul 2022 16:01:48	U	SMPL
CCB36 - 1	CCB36	10	-0.0752	-	245.1	PPB	14 Jul 2022 16:04:03	U	SMPL
CRA - 1	CRA	483	0.0517	-	245.1	PPB	14 Jul 2022 16:06:18	U	SMPL
HighStd - 1	HighStd	37527	9.9960	-	245.1	PPB	14 Jul 2022 16:08:34	U	SMPL
ChkStd - 1	ChkStd	26536	7.0455	-	245.1	PPB	14 Jul 2022 16:10:50	U	SMPL
PB146233BL - 1	PBW	7	-0.0760	-	245.1	PPB	14 Jul 2022 16:13:07	U	SMPL
PB146233BS - 1	LCSW	12972	3.4044	-	245.1	PPB	14 Jul 2022 16:15:23	U	SMPL
N3214-10 - 1	LOD-MDL-WATER-04-QT2-2022	452	0.0434	-	245.1	PPB	14 Jul 2022 16:17:40	U	SMPL
N3214-11 - 1	LOQ-WATER-05-QT2-2022	985	0.1865	-	245.1	PPB	14 Jul 2022 16:22:53	U	SMPL
CCV37 - 1	CCV37	17750	4.6870	-	245.1	PPB	14 Jul 2022 16:25:09	U	SMPL
CCB37 - 1	CCB37	23	-0.0717	-	245.1	PPB	14 Jul 2022 16:27:26	U	SMPL