

LB121028


7470A

INSTRUMENT ID: C114

Linear

 μ Abs.:

41028



A= 0.0000e+000

B= 2.3759e-004

C= 1.7456e-002

Rho= 0.9989940

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	O/D
0.0	0.000	0.029	0.029	47	0.000	47					-
0.2	0.200	0.182	-0.018	691	0.0 %	691					-9
2.5	2.500	2.374	-0.126	9918	0.0 %	9918					-5
5.0	5.000	5.051	0.051	21186	0.0 %	21186					1
7.5	7.500	7.799	0.299	32753	0.0 %	32753					4
10.0	10.000	9.765	-0.235	41028	0.0 %	41028					-2

LB121086 INSTRUMENT ID :: CV1

Method: 7470A		Operator: Admin		Date of Analysis: 15 Jul 2022 10:20:44					
Sample ID	Extended ID	µ Abs.	Conc.	Std Cond	Method	Units	Date	Type	Type
0.0 - 1	50	47	-	0.0000	7470A	PPB	15 Jul 2022 10:31:43	S	Std
0.2 - 1	50.2	691	-	0.2000	7470A	PPB	15 Jul 2022 10:33:59	S	Std
2.5 - 1	52.5	9918	-	2.5000	7470A	PPB	15 Jul 2022 10:36:16	S	Std
5.0 - 1	55	21186	-	5.0000	7470A	PPB	15 Jul 2022 10:38:33	S	Std
7.5 - 1	57.5	32753	-	7.5000	7470A	PPB	15 Jul 2022 10:40:50	S	Std
10.0 - 1	510	41028	-	10.0000	7470A	PPB	15 Jul 2022 10:43:05	S	Std
ICV41 - 1	ICV41	16627	3.9679	-7470A		PPB	15 Jul 2022 10:45:54	U	SMPL
ICB41 - 1	ICB41	-22	0.0122	-7470A		PPB	15 Jul 2022 10:48:08	U	SMPL
CCV38 - 1	CCV38	20020	4.7741	-7470A		PPB	15 Jul 2022 10:53:11	U	SMPL
CCB38 - 1	CCB38	-11	0.0148	-7470A		PPB	15 Jul 2022 10:55:26	U	SMPL
CRA - 1	CRA	614	0.1633	-7470A		PPB	15 Jul 2022 10:57:40	U	SMPL
HighStd - 1	HighStd	39482	9.3981	-7470A		PPB	15 Jul 2022 10:59:55	U	SMPL
ChkStd - 1	ChkStd	28399	6.7649	-7470A		PPB	15 Jul 2022 11:04:51	U	SMPL
PB146229BL - 1	PBW	-112	-0.0092	-7470A		PPB	15 Jul 2022 11:07:05	U	SMPL
PB146229BS - 1	LCSW	15087	3.6020	-7470A		PPB	15 Jul 2022 11:09:22	U	SMPL
N3682-01 - 1	17177	851	0.2196	-7470A		PPB	15 Jul 2022 11:11:38	U	SMPL
N3682-02 - 1	17171	-149	-0.0179	-7470A		PPB	15 Jul 2022 11:13:55	U	SMPL
N3682-03 - 1	COMP-1-0712	-56	0.0042	-7470A		PPB	15 Jul 2022 11:16:11	U	SMPL
N3682-04 - 1	17166-OIL	-17	0.0134	-7470A		PPB	15 Jul 2022 11:18:28	U	SMPL
N3687-05 - 1	D2584	-38	0.0084	-7470A		PPB	15 Jul 2022 11:20:43	U	SMPL
CCV39 - 1	CCV39	20169	4.8095	-7470A		PPB	15 Jul 2022 11:22:58	U	SMPL
CCB39 - 1	CCB39	-28	0.0108	-7470A		PPB	15 Jul 2022 11:25:13	U	SMPL
N3687-06 - 1	D2585	-28	0.0108	-7470A		PPB	15 Jul 2022 11:27:29	U	SMPL
N3711-02 - 1	PSC124019	0	0.0175	-7470A		PPB	15 Jul 2022 11:29:44	U	SMPL
N3711-04 - 1	PSC124042	-20	0.0127	-7470A		PPB	15 Jul 2022 11:31:59	U	SMPL
N3711-06 - 1	PSC124035	942	0.2413	-7470A		PPB	15 Jul 2022 11:34:15	U	SMPL
N3711-08 - 1	PSC124055	-207	-0.0317	-7470A		PPB	15 Jul 2022 11:36:31	U	SMPL
N3711-08DUP - 1	PSC124055DUP	-50	0.0056	-7470A		PPB	15 Jul 2022 11:38:47	U	SMPL
N3711-08MS - 1	PSC124055MS	14790	3.5315	-7470A		PPB	15 Jul 2022 11:41:03	U	SMPL
N3711-08MSD - 1	PSC124055MSD	16102	3.8432	-7470A		PPB	15 Jul 2022 11:43:20	U	SMPL
PB146230BL - 1	PBW	-79	-0.0013	-7470A		PPB	15 Jul 2022 11:45:37	U	SMPL
PB146230BS - 1	LCSW	14644	3.4968	-7470A		PPB	15 Jul 2022 11:47:53	U	SMPL
CCV40 - 1	CCV40	19768	4.7142	-7470A		PPB	15 Jul 2022 11:50:08	U	SMPL
CCB40 - 1	CCB40	-73	0.0001	-7470A		PPB	15 Jul 2022 11:52:25	U	SMPL
N3682-05 - 1	17172	8	0.0194	-7470A		PPB	15 Jul 2022 11:54:41	U	SMPL
N3682-06 - 1	17166	15	0.0210	-7470A		PPB	15 Jul 2022 11:56:57	U	SMPL
N3697-01 - 1	PIPELINE-PIGGING-TK2045L	81	0.0367	-7470A		PPB	15 Jul 2022 11:59:12	U	SMPL
N3697-02 - 1	CASING-NUMBER-5D-WATERING	0	0.0175	-7470A		PPB	15 Jul 2022 12:01:30	U	SMPL
N3697-02DUP - 1	CASING-NUMBER-5D-WATERINGDUP	0	0.0175	-7470A		PPB	15 Jul 2022 12:03:46	U	SMPL
N3697-02MS - 1	CASING-NUMBER-5D-WATERINGMS	17381	4.1471	-7470A		PPB	15 Jul 2022 12:06:02	U	SMPL
N3697-02MSD - 1	CASING-NUMBER-5D-WATERINGMSD	16181	3.8619	-7470A		PPB	15 Jul 2022 12:08:19	U	SMPL
PB146201TB - 1	PB146201TB	-56	0.0042	-7470A		PPB	15 Jul 2022 12:10:35	U	SMPL
PB146202TB - 1	PB146202TB	-4	0.0165	-7470A		PPB	15 Jul 2022 12:12:52	U	SMPL
CCV41 - 1	CCV41	21126	5.0368	-7470A		PPB	15 Jul 2022 12:15:08	U	SMPL
CCB41 - 1	CCB41	-43	0.0072	-7470A		PPB	15 Jul 2022 12:17:24	U	SMPL
N3711-08LX5 - 1		-56	0.0042	-7470A		PPB	15 Jul 2022 12:19:39	U	SMPL
N3711-08A - 1		16959	4.0468	-7470A		PPB	15 Jul 2022 12:21:55	U	SMPL
N3697-02LX5 - 1		-106	-0.0077	-7470A		PPB	15 Jul 2022 12:24:10	U	SMPL
N3697-02A - 1		16669	3.9779	-7470A		PPB	15 Jul 2022 12:26:26	U	SMPL
CCV42 - 1	CCV42	21967	5.2367	-7470A		PPB	15 Jul 2022 12:28:41	U	SMPL
CCB42 - 1	CCB42	-99	-0.0061	-7470A		PPB	15 Jul 2022 12:30:57	U	SMPL