

LB 121101


7471B

INSTRUMENT ID: CV2

Linear

 μ Abs.:

54313



A= 0.0000e+000

B= 1.8376e-004

C= 5.8831e-002

Rho= 0.9999425

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.058	0.058	-3	0.000	-3					1
0.2	0.200	0.204	0.004	789	0.0 %	789					2
2.5	2.500	2.450	-0.050	13014	0.0 %	13014					2
5.0	5.000	4.957	-0.043	26653	0.0 %	26653					1
7.5	7.500	7.491	-0.009	40447	0.0 %	40447					0
10.0	10.000	10.039	0.039	54313	0.0 %	54313					0

LB121101 INSTRUMENT ID : CV1

Method: 7471B		Operator: Admin		Date of Analysis: 15 Jul 2022 13:24:39					
Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-3	-	0.0000	7471B	PPB	15 Jul 2022 13:28:52	S	Std
0.2 - 1	50.2	789	-	0.2000	7471B	PPB	15 Jul 2022 13:31:09	S	Std
2.5 - 1	52.5	13014	-	2.5000	7471B	PPB	15 Jul 2022 13:33:26	S	Std
5.0 - 1	55	26653	-	5.0000	7471B	PPB	15 Jul 2022 13:35:43	S	Std
7.5 - 1	57.5	40447	-	7.5000	7471B	PPB	15 Jul 2022 13:38:00	S	Std
10.0 - 1	510	54313	-	10.0000	7471B	PPB	15 Jul 2022 13:40:17	S	Std
ICV42 - 1	ICV42	21949	4.0922	-	7471B	PPB	15 Jul 2022 13:43:34	U	SMPL
ICB42 - 1	ICB42	-53	0.0491	-	7471B	PPB	15 Jul 2022 13:45:49	U	SMPL
CCV43 - 1	CCV43	26412	4.9124	-	7471B	PPB	15 Jul 2022 13:53:16	U	SMPL
CCB43 - 1	CCB43	-54	0.0489	-	7471B	PPB	15 Jul 2022 13:55:30	U	SMPL
CRA - 1	CRA	786	0.2033	-	7471B	PPB	15 Jul 2022 13:57:46	U	SMPL
HighStd - 1	HighStd	52146	9.6413	-	7471B	PPB	15 Jul 2022 14:00:01	U	SMPL
ChkStd - 1	ChkStd	38305	7.0978	-	7471B	PPB	15 Jul 2022 14:02:17	U	SMPL
PB146255BL - 1	PBS	-126	0.0357	-	7471B	PPB	15 Jul 2022 14:04:34	U	SMPL
PB146255BS - 1	LCSS	18596	3.4761	-	7471B	PPB	15 Jul 2022 14:06:52	U	SMPL
N3693-02 - 1	VNJ-230	2210	0.4649	-	7471B	PPB	15 Jul 2022 14:09:08	U	SMPL
N3693-02DUP - 1	VNJ-230DUP	2293	0.4802	-	7471B	PPB	15 Jul 2022 14:11:25	U	SMPL
N3693-02MS - 1	VNJ-230MS	23412	4.3611	-	7471B	PPB	15 Jul 2022 14:13:42	U	SMPL
N3693-02MSD - 1	VNJ-230MSD	22853	4.2583	-	7471B	PPB	15 Jul 2022 14:15:59	U	SMPL
N3708-01 - 1	OR-2-071322	1818	0.3929	-	7471B	PPB	15 Jul 2022 14:18:15	U	SMPL
CCV44 - 1	CCV44	26880	4.9984	-	7471B	PPB	15 Jul 2022 14:20:31	U	SMPL
CCB44 - 1	CCB44	-37	0.0520	-	7471B	PPB	15 Jul 2022 14:22:47	U	SMPL
N3711-01 - 1	PSC124019	3893	0.7742	-	7471B	PPB	15 Jul 2022 14:25:04	U	SMPL
N3711-03 - 1	PSC124042	4383	0.8643	-	7471B	PPB	15 Jul 2022 14:27:19	U	SMPL
N3711-05 - 1	PSC124035	2121	0.4486	-	7471B	PPB	15 Jul 2022 14:29:35	U	SMPL
N3711-07 - 1	PSC124055	8143	1.5552	-	7471B	PPB	15 Jul 2022 14:31:51	U	SMPL
N3716-01 - 1	COMP-1	5123	1.0002	-	7471B	PPB	15 Jul 2022 14:34:07	U	SMPL
N3716-03 - 1	VNJ-236	1885	0.4052	-	7471B	PPB	15 Jul 2022 14:36:23	U	SMPL
N3716-05 - 1	ETGI-290	35544	6.5905	-	7471B	PPB	15 Jul 2022 14:38:39	U	SMPL
N3720-01 - 1	RIDGEFIELD-MR-SP	1566	0.3466	-	7471B	PPB	15 Jul 2022 14:40:56	U	SMPL
N3724-01 - 1	OE-1-071422	9862	1.8711	-	7471B	PPB	15 Jul 2022 14:43:18	U	SMPL
N3731-01 - 1	TP-1	573	0.1641	-	7471B	PPB	15 Jul 2022 14:45:33	U	SMPL
CCV45 - 1	CCV45	26473	4.9236	-	7471B	PPB	15 Jul 2022 14:47:49	U	SMPL
CCB45 - 1	CCB45	70	0.0717	-	7471B	PPB	15 Jul 2022 14:50:04	U	SMPL
N3731-03 - 1	TP-2	26284	4.8888	-	7471B	PPB	15 Jul 2022 14:52:21	U	SMPL
N3693-02LX5 - 1		426	0.1371	-	7471B	PPB	15 Jul 2022 14:54:37	U	SMPL
N3693-02A - 1		23688	4.4118	-	7471B	PPB	15 Jul 2022 14:56:54	U	SMPL
CCV46 - 1	CCV46	27337	5.0823	-	7471B	PPB	15 Jul 2022 14:59:09	U	SMPL
CCB46 - 1	CCB46	-32	0.0530	-	7471B	PPB	15 Jul 2022 15:01:27	U	SMPL